


NO: SAMM 979

Page: 1 of 1

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	Quality Control Laboratory, Duopharma (M) Sendirian Berhad Lot 2599, Jalan Seruling 59 Kawasan 3 Taman Klang Jaya 41200 Klang, Selangor , 41200, SELANGOR MALAYSIA
	
<b>ACCREDITED SINCE :</b>	13 OCTOBER 2025
<b>FIELD(S) OF TESTING:</b>	MICROBIOLOGICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

<b>CENTRAL LOCATION:</b>	Quality Control Laboratory, Duopharma (M) Sendirian Berhad Lot 2599, Jalan Seruling 59 Kawasan 3 Taman Klang Jaya 41200 Klang, Selangor , 41200, Selangor
<b>FIELD(S) OF TESTING :</b>	MICROBIOLOGICAL,

**SCOPE OF TESTING : MICROBIOLOGICAL**

<b>Material / Product Tested</b>	<b>Type Of Test / Properties Measured / Range Of Measurement</b>	<b>Standard Test Methods / Equipment / Techniques</b>
<b>Pharmaceutical Product:</b> (haemodialysis Solutions)	Bioburden Test:  1. Microbial Enumeration 2. Membrane Filtration  Bacterial Endotoxins Test	QCSC.010.14 based on USP 43 <61>  QCSC.049.11 based on USP 43 <85> Kinetic Chromogenic Method
<b>Pharmaceutical Product:</b> (eye Drops)	Sterility Tests  Sterility Tests (Biological Indicator)	QCSC.048.18 based on USP 43 <71>