

# Schedule

Issue date: 30 January 2026  
Valid Until: 30 January 2031



NO: SAMM 859

Page: 1 of 2

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>  	BS METROLOGY AND SERVICES 18, LORONG PERMAI JAYA 13, TAMAN PERMAI JAYA, 14000 BUKIT MERTA JAM, , 14000, PULAU PINANG MALAYSIA
<b>ACCREDITED SINCE :</b>	30 JANUARY 2026
<b>FIELD(S) OF CALIBRATION:</b>	Surface Plate Flatness DIMENSIONAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

**\* The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

<b>CENTRAL LOCATION</b>	BS METROLOGY AND SERVICES 18, LORONG PERMAI JAYA 13, TAMAN PERMAI JAYA, 14000 BUKIT MERTA JAM, , 14000, Pulau Pinang
<b>FIELD(S) OF CALIBRATION :</b>	, DIMENSIONAL

## SCOPE OF CALIBRATION : Surface Plate Flatness

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30004503> for the current scope of accreditation

# Schedule

Issue date: 30 January 2026  
Valid Until: 30 January 2031



NO: SAMM 859

Page: 2 of 2

## SCOPE OF CALIBRATION : DIMENSIONAL

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
Surface Plate Flatness	(300 x 300) mm up to (2000 x 1500) mm	2.3 $\mu$ m	Calibrated Using Planekator, Micro-Indicator with reference to JIS B 7513:1992 & BS817:2008.

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30004503> for the current scope of accreditation