


# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 1 of 12

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	Toyotech Engineering Sdn. Bhd. 2G, Block E, Jalan Perubatan 3 Pandan Indah 55100 WP Kuala Lumpur , 55100, WILAYAH PERSEKUTUAN KUALA LUMPUR MALAYSIA
	
<b>ACCREDITED SINCE :</b>	18 MARCH 2026
<b>FIELD(S) OF CALIBRATION:</b>	ELECTRICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

**\* The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

<b>CENTRAL LOCATION</b>	Toyotech Engineering Sdn. Bhd. 2G, Block E, Jalan Perubatan 3 Pandan Indah 55100 WP Kuala Lumpur , 55100, Wilayah Persekutuan Kuala Lumpur
<b>FIELD(S) OF CALIBRATION :</b>	ELECTRICAL,

## SCOPE OF CALIBRATION : ELECTRICAL

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
indicating Instruments Dc Voltage	0 mV to 220 mV	0.4 $\mu$ V	Generating using Calibrator model Fluke 5700A
	221 mV to 1100 V	2 $\mu$ V/V	
indicating Instruments Dc Current	0 $\mu$ A to 220 $\mu$ A	2 nA	Generating using Calibrator model Fluke 5700A
	221 $\mu$ A to 220 mA	13 nA	
	221 mA to 2.2 A	73 $\mu$ A	Generating using Calibrator model Fluke 5500A
	2.3 A to 11 A	2 mA	

# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 2 of 12

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
<b>indicating Instruments</b> Ac Voltage	<u>1 mV to 22 mV</u> 45 Hz to 1 kHz <u>23 mV to 220 mV</u> 45 Hz to 1 MHz <u>221 mV to 2.2 V</u> 45 Hz to 1 MHz <u>2.3 V to 22 V</u> 45 Hz to 1 MHz <u>23 V to 220 V</u> 45 Hz to 100 kHz <u>221 V to 1100 V</u> 45 Hz to 10 kHz	<u>3 <math>\mu</math>V</u> <u>7 <math>\mu</math>V</u> <u>39 <math>\mu</math>V</u> <u>440 <math>\mu</math>V</u> <u>5 mV</u> <u>33 mV</u>	Generating using Calibrator model Fluke 5700A
<b>indicating Instruments</b> Ac Current	<u>10 <math>\mu</math>A to 220 <math>\mu</math>A</u> 45 Hz to 10 kHz <u>0.23 mA to 2.2 mA</u> 45 Hz to 10 kHz <u>2.3 mA to 22 mA</u> 45 Hz to 10 kHz <u>23 mA to 220 mA</u> 45 Hz to 10 kHz <u>0.23 A to 2.2 A</u> 45 Hz to 10 kHz <u>2.2 A to 11 A</u> 45 Hz to 10 kHz	<u>5 nA</u> <u>150 nA</u> <u>2 <math>\mu</math>A</u> <u>18 <math>\mu</math>A</u> <u>210 <math>\mu</math>A</u> <u>5 mA</u>	Generating using Calibrator model Fluke 5500A
<b>indicating Instruments</b> Ac Current At 60 Hz	11 A to 50 A	85 mA	Generating using AC Voltage Current Standard model Yokogawa 2558
<b>indicating Instruments</b> Dc Resistance	<u>0 <math>\Omega</math> to 10 <math>\Omega</math></u> <u>11 <math>\Omega</math> to 100 k<math>\Omega</math></u> <u>101 k<math>\Omega</math> to 1 M<math>\Omega</math></u> <u>1.1 M<math>\Omega</math> to 10 M<math>\Omega</math></u> <u>11 M<math>\Omega</math> to 100 M<math>\Omega</math></u>	<u>0.1 m<math>\Omega</math></u> <u>3 m<math>\Omega</math></u> <u>7 <math>\Omega</math></u> <u>89 <math>\Omega</math></u> <u>4 k<math>\Omega</math></u>	Generating using Calibrator model Fluke 5700A
<b>indicating Instruments</b> Insulation Resistance	<u>0.1 <math>\Omega</math> to 1 <math>\Omega</math></u> <u>1.1 <math>\Omega</math> to 10 <math>\Omega</math></u> <u>10.1 <math>\Omega</math> to 100 <math>\Omega</math></u> <u>100.1 <math>\Omega</math> to 1000 <math>\Omega</math></u> <u>1000.1 <math>\Omega</math> to 10 k<math>\Omega</math></u> <u>10.1 k<math>\Omega</math> to 100 k<math>\Omega</math></u> <u>100.1 k<math>\Omega</math> to 1 M<math>\Omega</math></u> <u>1.1 M<math>\Omega</math> to 10 M<math>\Omega</math></u> <u>10.1 M<math>\Omega</math> to 110 M<math>\Omega</math></u> <u>111 M<math>\Omega</math> to 1 G<math>\Omega</math></u> <u>1 G<math>\Omega</math> to 10 G<math>\Omega</math></u>	<u>0.5 m<math>\Omega</math></u> <u>0.6 m<math>\Omega</math></u> <u>1 m<math>\Omega</math></u> <u>27 m<math>\Omega</math></u> <u>0.1 <math>\Omega</math></u> <u>1 <math>\Omega</math></u> <u>14 <math>\Omega</math></u> <u>0.6 k<math>\Omega</math></u> <u>36 k<math>\Omega</math></u> <u>53 k<math>\Omega</math></u> <u>30 M<math>\Omega</math></u>	Generating using Decade Resistance Box Yokogawa 2793-01 and 2793-03

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30006235> for the current scope of accreditation

# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 3 of 12

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
<b>indicating Instruments</b> Temperature Indicator Electrical R-type K-type E-type J-type T-type (without Cold Junction Compensation)	0 to 1600 - $^{\circ}$ C -200 to 1200 - $^{\circ}$ C 0 to 700 - $^{\circ}$ C -200 to 600 - $^{\circ}$ C -200 to 200 - $^{\circ}$	0.3 - $^{\circ}$ C 0.4 - $^{\circ}$ C 0.2 - $^{\circ}$ C 0.2 - $^{\circ}$ C 0.1 - $^{\circ}$ C	Electrical Simulation using DC Voltage Standard (refer to ITS-90)
<b>indicating Instruments</b> Frequency	1 Hz to 10 kHz 11 kHz to 100 kHz 101 kHz to 1 MHz 1.1 MHz to 10 MHz 11 MHz to 100 MHz 101 MHz to 1 GHz 1.1 GHz to 2 GHz 2.1 GHz to 3 GHz	8 mHz 36 mHz 2 mHz 18 mHz 0.2 Hz 2 Hz 9 Hz 7 Hz	Generating using Function Generator Yokogawa FG-120 Generating using Signal Generator Agilent E4436B
<b>indicating Instruments</b> Rf Power	<u>100 kHz to 3 GHz</u> 10 dBm to - 120 dBm	0.3 dB	Generating using Signal Generator Agilent E4436B
<b>indicating Instruments</b> Frequency Modulation Deviation	<u>Carrier Frequency:</u> <u>100 kHz to 3 GHz @</u> <u>Internal</u> <u>Rate : 1 Hz to 500 kHz</u>  0.5 kHz to 1 kHz 1.1 kHz to 10 kHz 10.1 kHz to 100 kHz	20 Hz 200 Hz 2 kHz	Generating using Signal Generator Agilent E4436B
<b>indicating Instruments</b> Amplitude Modulation Depth	Carrier Frequency: 100 kHz to 3 GHz @ Internal Rate : 1Hz to 500 kHz  0 % to 30 % 31 % to 60 % 61 % to 100 %	1 % 2 % 4 %	Generating using Signal Generator Agilent E4436B
<b>indicating Instruments</b> Distortion	<u>20 Hz to 20 kHz</u> -10 dB -20 dB -30 dB to -60 dB	0.5 dB 0.6 dB 0.8 dB	Generating using Distortion Calibrator Meguro MKS-682
<b>sources</b> Dc Voltage	0 mV to 100 mV 101 mV to 1 V 1.1 V to 10 V 10.1 V to 100V 100.1 V to 1000 V	0.6 $\mu$ V 8 $\mu$ V 45 $\mu$ V 0.7 mV 7 mV	Measurement using Multimeter Agilent 3458A

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30006235> for the current scope of accreditation

# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 4 of 12

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
<b>sources</b> Dc Current	0 mA to 10 mA	17 $\mu$ A	Measurement using Multimeter Agilent 3458A
	10.1 mA to 100 mA	2 $\mu$ A	
	100.1 mA to 1 A	54 mA	Measurement using Multimeter Agilent 34401A
	1.1 A to 3 A	0.6 mA	
<b>sources</b> Ac Voltage	1 mV to 10 mV	4 $\mu$ V	Measurement using Multimeter Agilent 3458A
	45 Hz to 1 kHz	4 $\mu$ V	
	1 mV to 100 mV	22 $\mu$ V	
	45 Hz to 1 MHz	0.3 mV	
	101 mV to 1 V	4 mV	
	45 Hz to 1 MHz	32 mV	
	1.1 V to 10 V		
	45 Hz to 1 MHz		
<b>sources</b> Ac Current	1 mA to 1 A	0.4 mA	Measurement using Multimeter Agilent 34401A
	45 Hz to 1 kHz	0.5 mA	
	1.1 A to 3 A		
	45 Hz to 1 kHz		
<b>sources</b> Dc Resistance	0 $\Omega$ to 10 $\mu\Omega$	0.3 m $\Omega$	Measurement using Multimeter Agilent 3458A
	10.1 $\mu\Omega$ to 100 $\Omega$	2 m $\Omega$	
	100.1 $\Omega$ to 1000 $\Omega$	17 m $\Omega$	
	1.1 k $\Omega$ to 10 k $\Omega$	180 m $\Omega$	
	10.1 k $\Omega$ to 100 k $\Omega$	2 $\Omega$	
	100.1 k $\Omega$ to 1 M $\Omega$	25 $\Omega$	
	1.1 M $\Omega$ to 10 M $\Omega$	0.7 k $\Omega$	
	10.1 M $\Omega$ to 100 M $\Omega$	13 k $\Omega$	
100.1 M $\Omega$ to 1 G $\Omega$	0.7 M $\Omega$		
<b>sources</b> Dc High Voltage	1 kV to 2 kV	6 V	Measurement using High Voltage Meter Kikusui 149-10A
	2.1 kV to 4 kV	7 V	
	4.1 kV to 6 kV	8 V	
	6.1 kV to 8 kV	8 V	
	8.1 kV to 10 kV	10 V	
<b>sources</b> Ac High Voltage At 50 Hz	1 kV to 2 kV	12 V	Measurement using High Voltage Meter Kikusui 149-10A
	2.1 kV to 4 kV	13 V	
	4.1 kV to 6 kV	23 V	
	6.1 kV to 8 kV	26 V	
	8.1 kV to 10 kV	29 V	

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30006235> for the current scope of accreditation

NO: SAMM 246

Page: 5 of 12

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
<b>sources</b> Leakage Dc Current	0.4 mA 0.5 mA 0.8 mA 1.0 mA 1.6 mA 2.0 mA 4.0 mA 5.0 mA	24 $\mu$ A 24 $\mu$ A 11 $\mu$ A 11 $\mu$ A 11 $\mu$ A 11 $\mu$ A 23 $\mu$ A 23 $\mu$ A	Measurement using Current Calibrator for W. Tester Kikusui TOS1200
<b>sources</b> Leakage Ac Current At 50 Hz	0.4 mA 0.5 mA 0.8 mA 1.0 mA 1.6 mA 2.0 mA 4.0 mA 5.0 mA 8.0 mA 10.0 mA 16.0 mA 20.0 mA 40.0 mA 50.0 mA	27 $\mu$ A 27 $\mu$ A 10 $\mu$ A 10 $\mu$ A 20 $\mu$ A 20 $\mu$ A 50 $\mu$ A 50 $\mu$ A 100 $\mu$ A 100 $\mu$ A 200 $\mu$ A 200 $\mu$ A 400 $\mu$ A 400 $\mu$ A	Measurement using Current Calibrator for W. Tester Kikusui TOS1200
<b>sources</b> Time Interval	1 s to 2 s 250 ms/div 2 s to 4 s (500 ms/div) 4 s to 8 s (1 s/div) 8 s to 20 s (2.5 s/div) 20 s to 60 s (5 s/div)	0.6 s 0.6 s 0.6 s 0.6 s 0.6 s	Measurement using Oscilloscope Tektronix TDS350
<b>sources</b> Time Base	10 MHz	0.3 Hz	Measurement using Frequency Counter Agilent 5335A
<b>sources</b> Frequency	10 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 100 MHz 100 Mhz to 1 GHz	50 $\mu$ Hz 110 $\mu$ Hz 0.4 mHz 4 mHz 66 mHz 360 mHz 3 Hz 27 Hz	Measurement using Frequency Counter Agilent 5385A

# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 6 of 12

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
	1 GHz to 2 GHz 2 GHz to 3 GHz	2 Hz 20 Hz	Measurement using Frequency Counter Anritsu MF76A
<b>sources</b> Distortion	10 Hz to 20 kHz -10 dB to -60 dB	0.1 dB	Measurement using Audio Analyzer Agilent 8903B
<b>sources</b> Attenuation	<u>10 Hz to 20 kHz</u> -10 dB to -60 dB	0.2 dB	Measurement using Audio Analyzer Agilent 8903B
<b>sources</b> Rf Power	<u>100 kHz to 1.3 GHz</u> 10 dBm to -120 dBm -20 dBm to -120 dBm	0.9 dB 1 dB	Measurement using Measuring Receiver Agilent 8902A
<b>sources</b> Frequency Modulation Deviation	<u>Carrier Frequency:</u> <u>100 kHz to 1.3 GHz @</u> <u>internal rate</u> 20 Hz to 200 Hz 0.5 kHz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz	11 Hz 110 Hz 1kHz	Measurement using Measuring Receiver Agilent 8902A
<b>sources</b> Amplitude Modulation Deviation	<u>Carrier Frequency:</u> <u>100 kHz to 1.3 GHz @</u> <u>internal rate</u> <u>20 Hz to 200 kHz</u> 0% to 30% 30% to 60% 60% to 90%	1% 2% 3%	Measurement using Measuring Receiver Agilent 8902A
<b>oscilloscope</b> Vertical Deflection	Amplitude (Peak-to-peak) 1 M $\Omega$ Load  200 $\mu$ V to 100 V	2.5 mV/V + 1 $\mu$	Generating using Calibration Generator Tektronix PG506
<b>oscilloscope</b> Time Mark	1 ns to 5 s	1 ms/s	Generating using Time Mark Generator Tektronix TG501
<b>oscilloscope</b> Frequency Range	50 kHz to 250 MHz	0.7 of the least significant displayed digit	Measurement using Levelled Sine Wave Generator Tektronix SG503
<b>oscilloscope</b> Amplitude Range	<u>X0.01 Range</u> 5 mV to 50 mV <u>X0.1 Range</u> 50 mV to 500 mV <u>X1 Range</u> 500 mV to 5 V	5 % of reading 4 % of reading 3 % of reading	Measurement using Levelled Sine Wave Generator Tektronix SG503

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30006235> for the current scope of accreditation

# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 7 of 12

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
Dc Current Via Shunt (1 M $\dot{I}$ ‰)	0 A to 30 A	1 mA/A	Measurement using Agilent 3458A through Current Shunt Yokogawa 2215

<b>SITE LOCATION (HQ)</b>	1. Toyotech Engineering Sdn. Bhd. 2G, Block E, Jalan Perubatan 3 Pandan Indah 55100 WP Kuala Lumpur, MALAYSIA
<b>FIELD(S) OF CALIBRATION :</b>	ELECTRICAL

### SCOPE OF CALIBRATION : ELECTRICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
<b>indicating Instruments</b> Dc Voltage	0 mV to 220 mV 221 mV to 1100 V	0.4 $\mu$ V 2 $\mu$ V/V	Generating using Calibrator model Fluke 5700A
<b>indicating Instruments</b> Dc Current	0 $\mu$ A to 220 $\mu$ A 221 $\mu$ A to 220 mA 221 mA to 2.2 A 2.3 A to 11 A	2 nA 13 nA 73 $\mu$ A 2 mA	Generating using Calibrator model Fluke 5700A Generating using Calibrator model Fluke 5500A
<b>indicating Instruments</b> Ac Voltage	1 mV to 22 mV 45 Hz to 1 kHz 23 mV to 220 mV 45 Hz to 1 MHz 221 mV to 2.2 V 45 Hz to 1 MHz 2.3 V to 22 V 45 Hz to 1 MHz 23 V to 220 V 45 Hz to 100 kHz 221 V to 1100 V 45 Hz to 10 kHz	3 $\mu$ V 7 $\mu$ V 39 $\mu$ V 440 $\mu$ V 5 mV 33 mV	Generating using Calibrator model Fluke 5700A

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30006235> for the current scope of accreditation

# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 8 of 12

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
indicating Instruments Ac Current	<u>10 <math>\mu</math>A to 220 <math>\mu</math>A</u> 45 Hz to 10 kHz <u>0.23 mA to 2.2 mA</u> 45 Hz to 10 kHz <u>2.3 mA to 22 mA</u> 45 Hz to 10 kHz <u>23 mA to 220 mA</u> 45 Hz to 10 kHz <u>0.23 A to 2.2 A</u> 45 Hz to 10 kHz <u>2.2 A to 11 A</u> 45 Hz to 10 kHz	5 nA 150 nA 2 $\mu$ A 18 $\mu$ A 210 $\mu$ A 5 mA	Generating using Calibrator model Fluke 5500A
indicating Instruments Ac Current At 60 Hz	11 A to 50 A	85 mA	Generating using AC Voltage Current Standard model Yokogawa 2558
indicating Instruments Dc Resistance	0 $\Omega$ to 10 $\Omega$ 11 $\Omega$ to 100 k $\Omega$ 101 k $\Omega$ to 1 M $\Omega$ 1.1 M $\Omega$ to 10 M $\Omega$ 11 M $\Omega$ to 100 M $\Omega$	0.1 m $\Omega$ 3 m $\Omega$ 7 $\Omega$ 89 $\Omega$ 4 k $\Omega$	Generating using Calibrator model Fluke 5700A
indicating Instruments Insulation Resistance	0.1 $\Omega$ to 1 $\Omega$ 1.1 $\Omega$ to 10 $\Omega$ 10.1 $\Omega$ to 100 $\Omega$ 100.1 $\Omega$ to 1000 $\Omega$ 1000.1 $\Omega$ to 10 k $\Omega$ 10.1 k $\Omega$ to 100 k $\Omega$ 100.1 k $\Omega$ to 1 M $\Omega$ 1.1 M $\Omega$ to 10 M $\Omega$ 10.1 M $\Omega$ to 110 M $\Omega$ 111 M $\Omega$ to 1 G $\Omega$ 1 G $\Omega$ to 10 G $\Omega$	0.5 m $\Omega$ 0.6 m $\Omega$ 1 m $\Omega$ 27 m $\Omega$ 0.1 $\Omega$ 1 $\Omega$ 14 $\Omega$ 0.6 k $\Omega$ 36 k $\Omega$ 53 k $\Omega$ 30 M $\Omega$	Generating using Decade Resistance Box Yokogawa 2793-01 and 2793-03
indicating Instruments Temperature Indicator Electrical R-type K-type E-type J-type T-type (without Cold Junction Compensation)	0 to 1600 $^{\circ}$ C -200 to 1200 $^{\circ}$ C 0 to 700 $^{\circ}$ C -200 to 600 $^{\circ}$ C -200 to 200 $^{\circ}$ C	0.3 $^{\circ}$ C 0.4 $^{\circ}$ C 0.2 $^{\circ}$ C 0.2 $^{\circ}$ C 0.1 $^{\circ}$ C	Electrical Simulation using DC Voltage Standard (refer to ITS-90)
indicating Instruments Frequency	1 Hz to 10 kHz 11 kHz to 100 kHz 101 kHz to 1 MHz	8 mHz 36 mHz 2 mHz	Generating using Function Generator Yokogawa FG-120

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30006235> for the current scope of accreditation

# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 9 of 12

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
	1.1 MHz to 10 MHz 11 MHz to 100 MHz 101 MHz to 1 GHz 1.1 GHz to 2 GHz 2.1 GHz to 3 GHz	18 mHz 0.2 Hz 2 Hz 9 Hz 7 Hz	Generating using Signal Generator Agilent E4436B
<b>indicating Instruments</b> Rf Power	100 kHz to 3 GHz 10 dBm to – 120 dBm	0.3 dB	Generating using Signal Generator Agilent E4436B
<b>indicating Instruments</b> Frequency Modulation Deviation	<u>Carrier Frequency:</u> 100 kHz to 3 GHz @ <u>Internal Rate : 1 Hz to 500 kHz</u>  0.5 kHz to 1 kHz 1.1 kHz to 10 kHz 10.1 kHz to 100 kHz	20 Hz 200 Hz 2 kHz	Generating using Signal Generator Agilent E4436B
<b>indicating Instruments</b> Amplitude Modulation Depth	<u>Carrier Frequency:</u> 100 kHz to 3 GHz @ <u>Internal Rate : 1Hz to 500 kHz</u>  0 % to 30 % 31 % to 60 % 61 % to 100 %	1 % 2 % 4 %	Generating using Signal Generator Agilent E4436B
<b>indicating Instruments</b> Distortion	<u>20 Hz to 20 kHz</u> -10 dB -20 dB -30 dB to -60 dB	0.5 dB 0.6 dB 0.8 dB	Generating using Distortion Calibrator Meguro MKS-682
<b>sources</b> Dc Voltage	0 mV to 100 mV 101 mV to 1 V 1.1 V to 10 V 10.1 V to 100V 100.1 V to 1000 V	0.6 μV 8 μV 45 μV 0.7 mV 7 mV	Measurement using Multimeter Agilent 3458A
<b>sources</b> Dc Current	0 mA to 10 mA 10.1 mA to 100 mA 100.1 mA to 1 A 1.1 A to 3 A	17 μA 2 μA 54 mA 0.6 mA	Measurement using Multimeter Agilent 3458A  Measurement using Multimeter Agilent 34401A

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30006235> for the current scope of accreditation

# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 10 of 12

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
<b>sources</b> Ac Voltage	<u>1 mV to 10 mV</u> 45 Hz to 1 kHz <u>1 mV to 100 mV</u> 45 Hz to 1 MHz <u>101 mV to 1 V</u> 45 Hz to 1 MHz <u>1.1 V to 10 V</u> 45 Hz to 1 MHz <u>11 V to 100 V</u> 45 Hz to 100 kHz <u>101 V to 1000 V</u> 45 Hz to 1 kHz	4 $\mu$ V 4 $\mu$ V 22 $\mu$ V 0.3 mV 4 mV 32 mV	Measurement using Multimeter Agilent 3458A
<b>sources</b> Ac Current	1 mA to 1 A 45 Hz to 1 kHz 1.1 A to 3 A 45 Hz to 1 kHz	0.4 mA 0.5 mA	Measurement using Multimeter Agilent 34401A
<b>sources</b> Dc Resistance	0 $\Omega$ to 10 $\mu\Omega$ 10.1 $\mu\Omega$ to 100 $\Omega$ 100.1 $\Omega$ to 1000 $\Omega$ 1.1 k $\Omega$ to 10 k $\Omega$ 10.1 k $\Omega$ to 100 k $\Omega$ 100.1 k $\Omega$ to 1 M $\Omega$ 1.1 M $\Omega$ to 10 M $\Omega$ 10.1 M $\Omega$ to 100 M $\Omega$ 100.1 M $\Omega$ to 1 G $\Omega$	0.3 m $\Omega$ 2 m $\Omega$ 17 m $\Omega$ 180 m $\Omega$ 2 $\Omega$ 25 $\Omega$ 0.7 k $\Omega$ 13 k $\Omega$ 0.7 M $\Omega$	Measurement using Multimeter Agilent 3458A
<b>sources</b> Dc High Voltage	1 kV to 2 kV 2.1 kV to 4 kV 4.1 kV to 6 kV 6.1 kV to 8 kV 8.1 kV to 10 kV	6 V 7 V 8 V 8 V 10 V	Measurement using High Voltage Meter Kikusui 149-10A
<b>sources</b> Ac High Voltage At 50 Hz	1 kV to 2 kV 2.1 kV to 4 kV 4.1 kV to 6 kV 6.1 kV to 8 kV 8.1 kV to 10 kV	12 V 13 V 23 V 26 V 29 V	Measurement using High Voltage Meter Kikusui 149-10A
<b>sources</b> Leakage Dc Current	0.4 mA 0.5 mA 0.8 mA 1.0 mA 1.6 mA 2.0 mA 4.0 mA 5.0 mA	24 $\mu$ A 24 $\mu$ A 11 $\mu$ A 11 $\mu$ A 11 $\mu$ A 11 $\mu$ A 23 $\mu$ A 23 $\mu$ A	Measurement using Current Calibrator for W. Tester Kikusui TOS1200

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30006235> for the current scope of accreditation

NO: SAMM 246

Page: 11 of 12

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
<b>sources</b> Leakage Ac Current At 50 Hz	0.4 mA 0.5 mA 0.8 mA 1.0 mA 1.6 mA 2.0 mA 4.0 mA 5.0 mA 8.0 mA 10.0 mA 16.0 mA 20.0 mA 40.0 mA 50.0 mA	27 $\mu$ A 27 $\mu$ A 10 $\mu$ A 10 $\mu$ A 20 $\mu$ A 20 $\mu$ A 50 $\mu$ A 50 $\mu$ A 100 $\mu$ A 100 $\mu$ A 200 $\mu$ A 200 $\mu$ A 400 $\mu$ A 400 $\mu$ A	Measurement using Current Calibrator for W. Tester Kikusui TOS1200
<b>sources</b> Time Interval	1 s to 2 s 250 ms/div 2 s to 4 s (500 ms/div) 4 s to 8 s (1 s/div) 8 s to 20 s (2.5 s/div) 20 s to 60 s (5 s/div)	0.6 s 0.6 s 0.6 s 0.6 s	Measurement using Oscilloscope Tektronix TDS350
<b>sources</b> Time Base	10 MHz	0.3 Hz	Measurement using Frequency Counter Agilent 5335A
<b>sources</b> Frequency	10 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 1 GHz 1 GHz to 2 GHz 2 GHz to 3 GHz	50 $\mu$ Hz 110 $\mu$ Hz 0.4 mHz 4 mHz 66 mHz 360 mHz 3 Hz 27 Hz 2 Hz 20 Hz	Measurement using Frequency Counter Agilent 5385A
<b>sources</b> Distortion	10 Hz to 20 kHz -10 dB to -60 dB	0.1 dB	Measurement using Audio Analyzer Agilent 8903B
<b>sources</b> Attenuation	10 Hz to 20 kHz -10 dB to -60 dB	0.2 dB	Measurement using Audio Analyzer Agilent 8903B

# Schedule

Issue date: 01 April 2026  
Valid Until: -



NO: SAMM 246

Page: 12 of 12

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
<b>sources</b> Rf Power	<u>100 kHz to 1.3 GHz</u> 10 dBm to -120 dBm -20 dBm to -120 dBm	0.9 dB 1 dB	Measurement using Measuring Receiver Agilent 8902A
<b>sources</b> Frequency Modulation Deviation	<u>Carrier Frequency:</u> <u>100 kHz to 1.3 GHz @</u> <u>internal rate</u> <u>20 Hz to 200 Hz</u> 0.5 kHz to 1 kHz 1 kHz to 10 khz 10 kHz to 100 kHz	11 Hz 110 Hz 1kHz	Measurement using Measuring Receiver Agilent 8902A
<b>sources</b> Amplitude Modulation Deviation	<u>Carrier Frequency:</u> <u>100 kHz to 1.3 GHz @</u> <u>internal rate</u> <u>20 Hz to 200 kHz</u> 0% to 30% 30% to 60% 60% to 90%	1% 2% 3%	Measurement using Measuring Receiver Agilent 8902A
<b>oscilloscope</b> Vertical Deflection	Amplitude (Peak-to-peak) 1 M <sub>Ω</sub> Load  200 μV to 100 V	2.5 mV/V + 1 μ	Generating using Calibration Generator Tektronix PG506
<b>oscilloscope</b> Time Mark	1 ns to 5 s	1 ms/s	Generating using Time Mark Generator Tektronix TG501
<b>oscilloscope</b> Frequency Range	50 kHz to 250 MHz	0.7 of the least significant displayed digit	Measurement using Levelled Sine Wave Generator Tektronix SG503
<b>oscilloscope</b> Amplitude Range	<u>X0.01 Range</u> mV to 50 mV <u>X0.1 Range</u> 50 mV to 500 mV <u>X1 Range</u> 500 mV to 5 V	5 % of reading 4 % of reading 3 % of reading	Measurement using Levelled Sine Wave Generator Tektronix SG503
Dc Current Via Shunt (1 M <sub>Ω</sub> )	0 A to 30 A	1 mA/A	Measurement using Agilent 3458A through Current Shunt Yokogawa 2215

Scan this QR Code or visit <https://accreditation.jsm.gov.my/public/listing/cab/samm-ct/30006235> for the current scope of accreditation