

NO: SAMM 837

Page: 1 of 12

LABORATORY LOCATION/ CENTRAL OFFICE: 	CENTRE FOR RESEARCH SERVICES Level 2, Research Management & Innovation Complex University of Malaya 50603 WP Kuala Lumpur , 50603, WILAYAH PERSEKUTUAN KUALA LUMPUR MALAYSIA
ACCREDITED SINCE :	10 NOVEMBER 2024
FIELD(S) OF TESTING:	NUCLEIC ACID
SITE:	
1 . SITE LABORATORY(HQ) :	INFRA ANALYSIS LABORATORY RESEARCH SERVICES DIVISION INSTITUTE OF RESEARCH MANAGEMENT & SERVICES, LEVEL 3, RESEARCH MANAGEMENT & INNOVATION COMPLEX UNIVERSITI MALAYA, MALAYSIA
FIELD(S) OF TESTING :	MECHANICAL
2 . SITE LABORATORY(HQ) :	ALGAE RESEARCH LABORATORY INSTITUTE OF OCEAN AND EARTH SCIENCES (IOES), B207, BLOCK B, LEVEL 2, ADVANCE STUDIES COMPLEX UNIVERSITI MALAYA, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL
3 . SITE LABORATORY(HQ) :	ANTIVIRAL RESEARCH LABORATORY (AVL) TROPICAL INFECTIOUS DISEASES RESEARCH & EDUCATION CENTRE (TIDREC), LEVEL 2, HIGH IMPACT RESEARCH (HIR) BUILDING, UNIVERSITI MALAYA,, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL
4 . SITE LABORATORY(HQ) :	VACCINE RESEARCH LABORATORY (VRL) TROPICAL INFECTIOUS DISEASES RESEARCH & EDUCATION CENTRE (TIDREC) , LEVEL 2, HIGH IMPACT RESEARCH (HIR) BUILDING, UNIVERSITI MALAYA,, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL
5 . SITE LABORATORY(HQ) :	SURVEILLANCE LABORATORY (SL 1) TROPICAL INFECTIOUS DISEASES RESEARCH & EDUCATION CENTRE (TIDREC), LEVEL 2, HIGH IMPACT RESEARCH (HIR) BUILDING, UNIVERSITI MALAYA,, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL
6 . SITE LABORATORY(HQ) :	SURVEILLANCE LABORATORY (SL 2) TROPICAL INFECTIOUS DISEASES RESEARCH & EDUCATION CENTRE (TIDREC), LEVEL 2, HIGH IMPACT RESEARCH (HIR) BUILDING, UNIVERSITI MALAYA,, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL
7 . SITE LABORATORY(HQ) :	PV SOLAR ENERGY INVERTER TESTING LABORATORY UM POWER ENERGY DEDICATED ADVANCED CENTRE

Schedule

Issue date: 24 July 2025
Valid Until: -



NO: SAMM 837

Page: 2 of 12

	(UMPEDAC), LEVEL G, WISMA R&D, UNIVERSITI MALAYA JALAN PANTAI BAHARU, MALAYSIA
FIELD(S) OF TESTING :	ELECTRICAL
8 . SITE LABORATORY(HQ) :	CATEGORY I
FIELD(S) OF TESTING :	ELECTRICAL
9 . SITE LABORATORY(HQ) :	OPTICAL CALIBRATION FACILITY PHOTONICS RESEARCH CENTRE (PRC), BLOCK A, DEPT. OF PHYSICS FACULTY OF SCIENCE UNIVERSITI MALAY, MALAYSIA
FIELD(S) OF CALIBRATION:	OPTICAL & PHOTOMETRIC

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

CENTRAL LOCATION:	CENTRE FOR RESEARCH SERVICES Level 2, Research Management & Innovation Complex University of Malaya 50603 WP Kuala Lumpur , 50603, Wilayah Persekutuan Kuala Lumpur
FIELD(S) OF TESTING :	NUCLEIC ACID,

SCOPE OF TESTING : NUCLEIC ACID

Schedule

Issue date: 24 July 2025
Valid Until: -



NO: SAMM 837

Page: 3 of 12

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Blood (serum/plasma) Or Cell Culture Materials, Materials, Synthetic Rna	Detection of dengue virus	In-house method UM/SL/TM-1 based on dengue virus detection and serotyping using DengueDetect® primers Multiplex RT-PCR In-house method UM/SL/TM-4 based on dengue virus detection using RealTime RT-PCR.
Oral And Respiratory Specimens Eg. Nasopharyngeal Swab (nps), Oropharyngeal Swab (ops), Nasal Swab (ns), Sputum, Deep Throat Saliva Or cell Culture Materials	Detection of SARS-Coronavirus-2	UM/SL/TM-5 based on SARS-CoV-2 detection using Real-Time RT-PCR

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3001022> for the current scope of accreditation

Schedule

Issue date: 24 July 2025
Valid Until: -



NO: SAMM 837

Page: 4 of 12

SITE LOCATION (HQ)	1. INFRA ANALYSIS LABORATORY RESEARCH SERVICES DIVISION INSTITUTE OF RESEARCH MANAGEMENT & SERVICES, LEVEL 3, RESEARCH MANAGEMENT & INNOVATION COMPLEX UNIVERSITI MALAYA, MALAYSIA
FIELD(S) OF TESTING :	MECHANICAL

SCOPE OF TESTING : MECHANICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Physical Catalysts And Catalyst Carriers	- Surface Area (BET)/ Mesopore (2nm-50nm)	ASTM D3663-20

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3001022> for the current scope of accreditation

NO: SAMM 837

Page: 5 of 12

SITE LOCATION (HQ)	2. ALGAE RESEARCH LABORATORY INSTITUTE OF OCEAN AND EARTH SCIENCES (IOES), B207, BLOCK B, LEVEL 2, ADVANCE STUDIES COMPLEX UNIVERSITI MALAYA, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL

SCOPE OF TESTING : BIOLOGICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Microalgae Culture Originating From Various Sources Including Fresh Water, Marine Water, Sediment, Soil, Epiphytic On Substrates, Sludge	Microalgae taxonomic identification	UM/ARL/TM-1 based on i) APHA (2017) Part 10900 Identification of Aquatic Organisms: C. Key for Identification of Freshwater Algae Common in Water Supplies and Polluted Waters. ii) AlgaeBase Website (http://www.algaebase.org)

Schedule

Issue date: 24 July 2025
Valid Until: -



NO: SAMM 837

Page: 6 of 12

SITE LOCATION (HQ)	3. ANTIVIRAL RESEARCH LABORATORY (AVL) TROPICAL INFECTIOUS DISEASES RESEARCH & EDUCATION CENTRE (TIDREC), LEVEL 2, HIGH IMPACT RESEARCH (HIR) BUILDING, UNIVERSITI MALAYA,, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL

SCOPE OF TESTING : BIOLOGICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Chemical Disinfectants And Antiseptics	Virus inactivation assay (Log10 reduction)	UM/AVL/TM-1 based on BS EN 14476:2013+A2:2019

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3001022> for the current scope of accreditation

Schedule

Issue date: 24 July 2025
Valid Until: -



NO: SAMM 837

Page: 7 of 12

SITE LOCATION (HQ)	4. VACCINE RESEARCH LABORATORY (VRL) TROPICAL INFECTIOUS DISEASES RESEARCH & EDUCATION CENTRE (TIDREC) , LEVEL 2, HIGH IMPACT RESEARCH (HIR) BUILDING, UNIVERSITI MALAYA,, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL

SCOPE OF TESTING : BIOLOGICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Chemical Disinfectants/ Antiseptics And Antimicrobial Equipment	Antimicrobial assay/ Percent reduction (%)	UM/VRL/TM-1 based on ASTM E2315- 16

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3001022> for the current scope of accreditation

NO: SAMM 837

Page: 8 of 12

SITE LOCATION (HQ)	5. SURVEILLANCE LABORATORY (SL 1) TROPICAL INFECTIOUS DISEASES RESEARCH & EDUCATION CENTRE (TIDREC), LEVEL 2, HIGH IMPACT RESEARCH (HIR) BUILDING, UNIVERSITI MALAYA,, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL

SCOPE OF TESTING : BIOLOGICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Blood (serum/plasma)	Detection of dengue antibodies and antigen	UM/SL/TM-2 based on Dengue NS1, IgM and IgG ELISA UM/SL/TM-3 based on Dengue NS1Ag and IgM+IgG Ab fluorescence immunoassay (FIA) Rapid Test

NO: SAMM 837

Page: 9 of 12

SITE LOCATION (HQ)	6. SURVEILLANCE LABORATORY (SL 2) TROPICAL INFECTIOUS DISEASES RESEARCH & EDUCATION CENTRE (TIDREC), LEVEL 2, HIGH IMPACT RESEARCH (HIR) BUILDING, UNIVERSITI MALAYA,, MALAYSIA
FIELD(S) OF TESTING :	BIOLOGICAL

SCOPE OF TESTING : BIOLOGICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Blood (serum/plasma)	Detection of dengue antibodies and antigen	UM/SL/TM-2 based on Dengue NS1, IgM and IgG ELISA UM/SL/TM-3 based on Dengue NS1Ag and IgM+IgG Ab fluorescence immunoassay (FIA) Rapid Test

NO: SAMM 837

Page: 10 of 12

SITE LOCATION (HQ)	7. PV SOLAR ENERGY INVERTER TESTING LABORATORY UM POWER ENERGY DEDICATED ADVANCED CENTRE (UMPEDAC), LEVEL G, WISMA R&D, UNIVERSITI MALAYA JALAN PANTAI BAHARU, MALAYSIA
FIELD(S) OF TESTING :	ELECTRICAL

SCOPE OF TESTING : ELECTRICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Photovoltaic Inverter 100 W To 10 Kw	1) Power factor 2) Current harmonics 3) Direct current injection 4) Under and over voltage 5) Under and over frequency 6) Response to utility recovery	IEC 61727:2004
	Voltage flicker	IEC 61000-3-3:2013
	Anti-islanding	IEC 62116:2014

Schedule

Issue date: 24 July 2025
Valid Until: -



NO: SAMM 837

Page: 11 of 12

SITE LOCATION (HQ)	8. CATEGORY I
FIELD(S) OF TESTING :	ELECTRICAL

SCOPE OF TESTING : ELECTRICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Photovoltaic Inverter 10 Kw To 10 Mw	1) Power factor 2) Current harmonics 3) Direct current injection 4) Voltage fluctuation 5) Anti-islanding	IEC 61727:2004 Distribution Code for Peninsular Malaysia, Sabah, and F.T Labuan (Amendments) 2017
	Voltage flicker	IEC 61000-3-11:2017 Distribution Code for Peninsular Malaysia, Sabah, and F.T Labuan (Amendments) 2017

SITE LOCATION (HQ)	1. OPTICAL CALIBRATION FACILITY PHOTONICS RESEARCH CENTRE (PRC) BLOCK A, DEPT. OF PHYSICS FACULTY OF SCIENCE UNIVERSITI MALAY, MALAYSIA
FIELD(S) OF CALIBRATION :	OPTICAL & PHOTOMETRIC

SCOPE OF CALIBRATION : OPTICAL & PHOTOMETRIC

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
Optical Time Domain Reflectometers (otdrs) (single Mode Fibre) Length Scale	5 km	0.23 m	Measured by comparison to single mode fibre physical standard. Wavelength: 1310 nm and 1550 nm UM/OCF/TM-1: Calibration Method for Calibration of Optical-Time Domain Reflectometer (OTDR) based on IEC 61746 -1:2009

Schedule

Issue date: 24 July 2025
Valid Until: -



NO: SAMM 837

Page: 12 of 12

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
	10 km	0.38 m	Measured by comparison to single mode fibre physical standard. Wavelength: 1310 nm and 1550 nm UM/OCF/TM-1: Calibration Method for Calibration of Optical-Time Domain Reflectometer (OTDR) based on IEC 61746 -1:2009

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3001022> for the current scope of accreditation