

# Schedule


Issue date: 08 May 2025  
Valid Until: -



## NO: SAMM 834

(Issue 2, 08 May 2025 replacement  
of SAMM 834 dated 08 May 2025)

Page: 1 of 1

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	George Kent (Malaysia) Berhad 1115, Jalan Puchong Taman Meranti Jaya 47120 Puchong, Selangor , 47120, SELANGOR MALAYSIA
	
<b>ACCREDITED SINCE :</b>	08 MAY 2025
<b>FIELD(S) OF TESTING:</b>	MECHANICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

<b>CENTRAL LOCATION:</b>	George Kent (Malaysia) Berhad 1115, Jalan Puchong Taman Meranti Jaya 47120 Puchong, Selangor , 47120, Selangor
<b>FIELD(S) OF TESTING :</b>	MECHANICAL,

## SCOPE OF TESTING : MECHANICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Cold Potable Water Meter - Dn15 & Dn 20	1. Determination of Intrinsic Errors (of Indication)	OIML R 49-2:2013, Clause 7.4 or ISO 4064-2:2014, Clause 7.4
	Flowrate range: 2 L/h to 6300 L/h	
	2. Static Pressure Test  Pressure range: 0.3 bar to 40 bar	OIML R 49-2:2013, Clause 7.3 or ISO 4064-2:2014, Clause 7.3

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3001027> for the current scope of accreditation