


NO: SAMM 1090(Issue 2, 17 June 2025 replacement
of SAMM 1090 dated 17 June 2025)

Page: 1 of 2

LABORATORY LOCATION/ CENTRAL OFFICE: 	Arkoweight (M) Sdn. Bhd. D2-1-20, Parklane OUG Jalan 1/152 Off Jalan Puchong 58200 Kuala Lumpur , 58200, WILAYAH PERSEKUTUAN KUALA LUMPUR MALAYSIA
ACCREDITED SINCE :	17 JUNE 2025

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

SITE LOCATION (HQ)	1.
FIELD(S) OF CALIBRATION :	MASS

SCOPE OF CALIBRATION : MASS

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
Electronic Balances And Weighing Scales	Up to 50 g Up to 200 g Up to 6000 g	0.18 mg 0.83 mg 0.13 g	1. Calibration using standard weights of OIML classes E2, F1 and M1.
	Up to 20 kg Up to 60 kg Up to 100 kg Up to 500 kg Up to 1000 kg	0.14 g 0.03 kg 0.07 kg 0.13 kg 0.36 kg	2. Calibration method with reference to Euramet cg 18 v4.0, OIML R 76-1 Ed. 2006, and UKAS LAB14 Ed. 2015

Schedule

Issue date: 17 June 2025
Valid Until: -



NO: SAMM 1090

(Issue 2, 17 June 2025 replacement
of SAMM 1090 dated 17 June 2025)

Page: 2 of 2

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
High Capacity Weighing Scales And Tank Scales.	Up to 2000 kg Up to 3000 kg Up to 4000 kg Up to 5000 kg	1.5 kg 3.3 kg 3.9 kg 4.4 kg	1. Calibration using standard weights and substitution loads. 2. Standard weights of OIML class M1 up to total nominal value of 1000 kg 3. Calibration method with reference to Euramet cg 18 v4.0, OIML R 76-1 Ed. 2006, and UKAS LAB14 Ed. 2015

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3001085> for the current scope of accreditation