

Schedule

Issue date: 13 June 2025
Valid Until: -



NO: SAMM 255

(Issue 5, 13 June 2025 replacement
of SAMM 255 dated 13 June 2025)

Page: 1 of 3

LABORATORY LOCATION/ CENTRAL OFFICE: 	IMCS Specialist Sdn Bhd 21-A, Jalan Sungai Nibong 14A/KU4, Rantau Panjang , 42100, SELANGOR MALAYSIA
ACCREDITED SINCE :	28 APRIL 2025
FIELD(S) OF CALIBRATION:	PRESSURE

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

*** The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

CENTRAL LOCATION	IMCS Specialist Sdn Bhd 21-A, Jalan Sungai Nibong 14A/KU4, Rantau Panjang , 42100, Selangor
FIELD(S) OF CALIBRATION :	PRESSURE,

SCOPE OF CALIBRATION : PRESSURE

--

Schedule

Issue date: 13 June 2025
Valid Until: -



NO: SAMM 255

(Issue 5, 13 June 2025 replacement
of SAMM 255 dated 13 June 2025)

Page: 2 of 3

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ($\pm$)*	Remarks
Pressure Transmitter (pneumatic)	-1000 mbar to 0 mbar	3 mbar	1. The calibration procedure covers the repeatability, hysteresis and instrument error. 2. Calibrations may be given in other units by conversion. 3. Calibrations method using pressure calibrator or pressure meter as reference.
Analogue Pressure Gauge (pneumatic)	-1000 mbar to 0 mbar	16 mbar	1. The calibration procedure covers the repeatability, hysteresis and instrument error. 2. Calibrations may be given in other units by conversion. 3. Calibrations method using pressure calibrator or pressure meter as reference.

SITE LOCATION (HQ)	1. CATEGORY I
FIELD(S) OF CALIBRATION :	PRESSURE

SCOPE OF CALIBRATION : PRESSURE

Schedule

Issue date: 13 June 2025
Valid Until: -



NO: SAMM 255

(Issue 5, 13 June 2025 replacement
of SAMM 255 dated 13 June 2025)

Page: 3 of 3

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
Pressure Transmitter (pneumatic)	-1000 mbar to 0 mbar	3 mbar	1. The calibration procedure covers the repeatability, hysteresis and instrument error. 2. Calibrations may be given in other units by conversion. 3. Calibrations method using pressure calibrator or pressure meter as reference.
Analogue Pressure Gauge (pneumatic)	-1000 mbar to 0 mbar	16 mbar	1. The calibration procedure covers the repeatability, hysteresis and instrument error. 2. Calibrations may be given in other units by conversion. 3. Calibrations method using pressure calibrator or pressure meter as reference.

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/lab/samm-ct/3002441> for the current scope of accreditation