Schedule

Issue date: 07 January 2025

Valid Until: -



NO: SAMM 986

Page: 1 of 3

LABORATORY LOCATION:	Meditech Laboratory Services Sdn. Bhd.	
(PERMANENT LABORATORY)	440, Ground Floor Jalan Haruan 4/3, Oakland Commercial Centre.	
	70300 Seremban, Negeri Sembilan. , 70300,	
	NEGERI SEMBILAN	
	MALAYSIA	
ACCREDITED SINCE :	07 JANUARY 2025	
FIELD(S) OF MEDICAL TESTING:	CHEMICAL PATHOLOGY	
	HAEMATOLOGY	
	MEDICAL MICROBIOLOGY	

The standard used for assessment of this laboratory is MS ISO 15189:2022 (ISO 15189:2022, IDT).

A medical laboratory's fulfilment of the requirements of ISO 15189 means the laboratory meets both the technical competence requirements and the management system requirements necessary for it to consistently deliver technically valid test results. The management system requirements in ISO 15189 are written in language relevant to a medical laboratory's operations. Medical laboratories that implement ISO 15189 operate generally in accordance with the principles of ISO 9001. (See Joint IAF-ILAC-ISO Communiquè, November 2021)

Schedule

Issue date: 07 January 2025

Valid Until: -



NO: SAMM 986

Page: 2 of 3

CENTRAL LOCATION	Meditech Laboratory Services Sdn. Bhd. 440, Ground Floor Jalan Haruan 4/3, Oakland Commercial Centre. 70300 Seremban, Negeri Sembilan., 70300, Negeri Sembilan
FIELD(S) OF MEDICAL TESTING:	CHEMICAL PATHOLOGY, HAEMATOLOGYMEDICAL
	MICROBIOLOGY

SCOPE OF MEDICAL TESTING: CHEMICAL PATHOLOGY

Specimen Tested	Type of Test/	Test Methods, Specifications/
	Properties Measured/	Equipment/Techniques Used
Urine	Opiates Screening	Test Strip, Cut-off: 300ng/mL
		Method / as documented in
		MLS/LAB/WI.1.0; Index 1.0
	Cannabinoids Screening	Test Strip, Cut-off: 50ng/mL
		Method / as documented in
		MLS/LAB/WI.1.0; Index 1.0
	Amphetamine Screening	Test Strip, Cut-off: 1000ng/mL
		Method / as documented in
		MLS/LAB/WI.1.0; Index 1.0
	Urine Feme	Test Strip, 10 tests parameter
		Method / Mission U500 Urine
		Analyzer as documented in
		MLS/LAB/WI.1.0; Index 2.0
		MLS/LAB/WI.1.0; Index 3.0
		MLS/LAB/WI.1.0; Index 6.0
		Microscopic Method / Microscope as documented in
		MLS/LAB.WI.1.0; Index 4.0
	Urine Pregnancy Test	Test Strip, Sensitivity: 25mIU/mL
		Method / as documented in
		MLS/LAB/WI.1.0; Index 5.0

SCOPE OF MEDICAL TESTING: HAEMATOLOGY

Specimen Tested	Type of Test/	Test Methods, Specifications/
	Properties Measured/	Equipment/Techniques Used
Whole Blood	ABO & Rh Blood Grouping	Test Tube Method / Centrifuge as
		documented in MLS/LAB/WI.3.0;
		Index 1.0
		MLS/LAB/WI.3.0; Index 2.0
		MLS/LAB/WI.3.0; Index 3.0

Schedule

Issue date: 07 January 2025

Valid Until: -



NO: SAMM 986

Page: 3 of 3

SCOPE OF MEDICAL TESTING: MEDICAL MICROBIOLOGY

Specimen Tested	Type of Test/	Test Methods, Specifications/
	Properties Measured/	Equipment/Techniques Used
Whole Blood	Malaria Parasite	ELISA Method / Microplate
		washer, Incubator 37°C and
		Reader as documented in
		MLS/LAB/WI.2.0; Index 1.0
		Blood Smear Thick and Thin
		Method as documented in
		MLS/LAB/WI.2.0; Index 2.0
Serum Or Plasma	HBsAg Screening	ELISA Method / Microplate
		washer and Reader, Incubator
		37°C as documented in
		MLS/LAB/WI.2.0; Index 3.0
	Anti-HIV 1&2 Screening	ELISA Method / Incubator 37°C,
		Microplate washer and Reader as
		documented in
		MLS/LAB/WI.2.0; Index 4.0
		Serodia HIV 1/2 Mix Method as
		documented in MLS/LAB/WI.2.0;
		Index 8.0
	Hepatitis C Screening (HCV)	ELISA Method / Incubator 37°C,
		Microplate washer and Reader as
		documented in MLS/LAB/WI.2.0;
		Index 5.0
	VDRL	RPR Carbon Antigen Method /
		Rotator as documented in
		MLS/LAB/WI.2.0; Index 6.0
	TPPA	Treponema pallidum particle
		Agglutination
		Method as documented in
		MLS/LAB/WI.2.0; Index 7.0

NOTE: