


Schedule

Issue date: 17 November 2025
Valid Until: -



NO: SAMM 986

Page: 1 of 3

LABORATORY LOCATION: (PERMANENT LABORATORY) 	Meditech Laboratory Services Sdn. Bhd. 440, Ground Floor Jalan Haruan 4/3, Oakland Commercial Centre. 70300 Seremban, Negeri Sembilan. , 70300, NEGERI SEMBILAN MALAYSIA
ACCREDITED SINCE :	07 JANUARY 2025
FIELD(S) OF MEDICAL TESTING :	CHEMICAL PATHOLOGY HAEMATOLOGY MEDICAL MICROBIOLOGY

The standard used for assessment of this laboratory is MS ISO 15189:2022 (ISO 15189:2022, IDT).

A medical laboratory's fulfilment of the requirements of ISO 15189 means the laboratory meets both the technical competence requirements and the management system requirements necessary for it to consistently deliver technically valid test results. The management system requirements in ISO 15189 are written in language relevant to a medical laboratory's operations. Medical laboratories that implement ISO 15189 operate generally in accordance with the principles of ISO 9001. (See Joint IAF-ILAC-ISO Communiqué, November 2021)

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-mt/3002617> for the current scope of accreditation

NO: SAMM 986

Page: 2 of 3

CENTRAL LOCATION	Meditech Laboratory Services Sdn. Bhd. 440, Ground Floor Jalan Haruan 4/3, Oakland Commercial Centre. 70300 Seremban, Negeri Sembilan. , 70300, Negeri Sembilan
FIELD(S) OF MEDICAL TESTING :	CHEMICAL PATHOLOGY, HAEMATOLOGY MEDICAL MICROBIOLOGY

SCOPE OF MEDICAL TESTING : CHEMICAL PATHOLOGY

Specimen Tested	Type of Test/ Properties Measured/	Test Methods, Specifications/ Equipment/Techniques Used
Urine	Opiates Screening	Test Strip, Cut-off: 300ng/mL Method / as documented in MLS/LAB/WI.1.0; Index 1.0
	Cannabinoids Screening	Test Strip, Cut-off: 50ng/mL Method / as documented in MLS/LAB/WI.1.0; Index 1.0
	Amphetamine Screening	Test Strip, Cut-off: 1000ng/mL Method / as documented in MLS/LAB/WI.1.0; Index 1.0
	Urine Feme	Test Strip, 10 tests parameter Method / Mission U500 Urine Analyzer as documented in MLS/LAB/WI.1.0; Index 2.0 MLS/LAB/WI.1.0; Index 3.0 MLS/LAB/WI.1.0; Index 6.0 Microscopic Method / Microscope as documented in MLS/LAB.WI.1.0; Index 4.0
	Urine Pregnancy Test	Test Strip, Sensitivity: 25mIU/mL Method / as documented in MLS/LAB/WI.1.0; Index 5.0

SCOPE OF MEDICAL TESTING : HAEMATOLOGY

Specimen Tested	Type of Test/ Properties Measured/	Test Methods, Specifications/ Equipment/Techniques Used
Whole Blood	ABO & Rh Blood Grouping	Test Tube Method / Centrifuge as documented in MLS/LAB/WI.3.0; Index 1.0 MLS/LAB/WI.3.0; Index 2.0 MLS/LAB/WI.3.0; Index 3.0

NO: SAMM 986

Page: 3 of 3

SCOPE OF MEDICAL TESTING : MEDICAL MICROBIOLOGY

Specimen Tested	Type of Test/ Properties Measured/	Test Methods, Specifications/ Equipment/Techniques Used
Whole Blood	Malaria Parasite	ELISA Method / Microplate washer, Incubator 37°C and Reader as documented in MLS/LAB/WI.2.0; Index 1.0
		Blood Smear Thick and Thin Method as documented in MLS/LAB/WI.2.0; Index 2.0
Serum Or Plasma	HBsAg Screening	ELISA Method / Microplate washer and Reader, Incubator 37°C as documented in MLS/LAB/WI.2.0; Index 3.0
	Anti-HIV 1&2 Screening	ELISA Method / Incubator 37°C, Microplate washer and Reader as documented in MLS/LAB/WI.2.0; Index 4.0
		Serodia HIV 1/2 Mix Method as documented in MLS/LAB/WI.2.0; Index 8.0
	Hepatitis C Screening (HCV)	ELISA Method / Incubator 37°C, Microplate washer and Reader as documented in MLS/LAB/WI.2.0; Index 5.0
	VDRL	RPR Carbon Antigen Method / Rotator as documented in MLS/LAB/WI.2.0; Index 6.0
	TPPA	Treponema pallidum particle Agglutination Method as documented in MLS/LAB/WI.2.0; Index 7.0

NOTE :