


NO: SAMM 975(Issue 2, 24 July 2025 replacement
of SAMM 975 dated 24 July 2025)

Page: 1 of 1

LABORATORY LOCATION/ CENTRAL OFFICE: 	RS-Simulation Asia Laboratory (RSAL), RS-Simulation Asia Sdn. Bhd. No. 15, Jalan MJ14, Taman Industri Meranti Jaya, 47120 Puchong, Selangor , 47120, SELANGOR MALAYSIA
ACCREDITED SINCE :	24 JULY 2025

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

SITE LOCATION (HQ)	1. CATEGORY III
FIELD(S) OF CALIBRATION :	HEAT & TEMPERATURE

SCOPE OF CALIBRATION : HEAT & TEMPERATURE

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
Temperature Climatic Test Chambers	-70 °C to -40 °C > -40 °C to 0 °C > 0 °C to 100 °C > 100 °C to 180 °C	0.3 °C 0.3 °C 0.3 °C 0.4 °C	Based on DKD-R 5-7 method A, B, or C
Relative Humidity Climatic Test Chambers	10 % rh to 30 % rh > 30 % rh to 60 % rh > 60 % rh to 95 % rh (10 °C to 65 °C)	2.2 % rh 2.4 % rh 2.4 % rh	Based on DKD-R 5-7 method A, B, or C