

Schedule

Issue date: 12 August 2025
Valid Until: -



NO: SAMM 407

(Issue 2, 12 August 2025 replacement of SAMM 407 dated 14 October 2025)

Page: 1 of 7

| | |
|---|---|
| LABORATORY LOCATION/ CENTRAL OFFICE: | Trescal (Malaysia) Sdn. Bhd., Perak No. 23-23A, Tasek Mutiara 2 Pusat Perdagangan Tasek Mutiara 31400 Ipoh, Perak , 31400, PERAK MALAYSIA |
|  | |
| ACCREDITED SINCE : | 12 AUGUST 2025 |
| FIELD(S) OF CALIBRATION: | HEAT & TEMPERATURE MASS PRESSURE DIMENSIONAL |

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

*** The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

| | |
|----------------------------------|---|
| CENTRAL LOCATION | Trescal (Malaysia) Sdn. Bhd., Perak No. 23-23A, Tasek Mutiara 2 Pusat Perdagangan Tasek Mutiara 31400 Ipoh, Perak , 31400, Perak |
| FIELD(S) OF CALIBRATION : | HEAT & TEMPERATURE, MASS, PRESSURE, DIMENSIONAL |

SCOPE OF CALIBRATION : HEAT & TEMPERATURE

| Instrument Calibrated/Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks |
|---|---|---|---|
| Temperature Sensor Pt 100 | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.2 °C 0.3 °C 0.5 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |

Schedule

Issue date: 12 August 2025
Valid Until: -



NO: SAMM 407

(Issue 2, 12 August 2025 replacement of SAMM 407 dated 14 October 2025)

Page: 2 of 7

| Instrument Calibrated/Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks |
|---|---|---|---|
| Temperature Sensor Thermocouple | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.2 °C 0.6 °C 0.7 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |
| Temperature Sensor Temperature Gauge | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.3 °C 0.5 °C 0.6 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |
| Temperature Sensor With Indicator Pt 100 | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.1 °C 0.3 °C 0.5 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |
| Temperature Sensor With Indicator Thermocouple | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.1 °C 0.6 °C 0.7 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |
| Temperature Measuring Device (by Electrical Simulation) A) Type K | -200 °C to -100 °C -100 °C to 1300 °C | 1.2 °C 0.2 °C | Calibrated by electrical simulation using calibrator and reference table ITS 90 |
| Temperature Measuring Device (by Electrical Simulation) B) Type J | -200 °C to 1200 °C | 0.3 °C | Calibrated by electrical simulation using calibrator and reference table ITS 90 |
| Temperature Measuring Device (by Electrical Simulation) C) Pt 100 | -200 °C to 850 °C | 0.2 °C | Calibrated by electrical simulation using calibrator and reference table ITS 90 |

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/ceb/samm-ct/3002737> for the current scope of accreditation

Schedule

Issue date: 12 August 2025
Valid Until: -



NO: SAMM 407

(Issue 2, 12 August 2025 replacement
of SAMM 407 dated 14 October 2025)

Page: 3 of 7

SCOPE OF CALIBRATION : MASS

| Instrument Calibrated/Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks |
|---|---|---|--|
| Standard Weight | 2 kg 5 kg 10 kg 20 kg 25 kg | 0.2 g 0.2 g 0.2 g 0.2 g 0.2 g | Calibrated by using standard weights and weighing comparator |

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3002737> for the current scope of accreditation

Schedule

Issue date: 12 August 2025
Valid Until: -



NO: SAMM 407

(Issue 2, 12 August 2025 replacement
of SAMM 407 dated 14 October 2025)

Page: 4 of 7

SCOPE OF CALIBRATION : PRESSURE

| Instrument Calibrated/Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks |
|--|--|---|--------------------------------|
| Pressure Measuring Device A) Hydraulic | 0 bar to 300 bar 300 bar to 700 bar | 0.07 bar 0.13 bar | Calibrated by using test gauge |
| Pressure Measuring Device B) Pneumatic | 0 bar to 0.005 bar 0.005 bar to 0.01 bar 0.01 bar to 0.03 bar 0.03 bar to 0.05 bar 0 bar to 30 bar | 0.004 mbar 0.006 mbar 0.02 mbar 0.04 mbar 0.005 bar | Calibrated by using test gauge |
| Pressure Measuring Device C) Vacuum | -0.9 bar to 0 bar | 0.002 bar | Calibrated by using test gauge |

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3002737> for the current scope of accreditation

NO: SAMM 407(Issue 2, 12 August 2025 replacement
of SAMM 407 dated 14 October 2025)

Page: 5 of 7

SCOPE OF CALIBRATION : DIMENSIONAL

| Instrument Calibrated/Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks |
|---|--|---|--|
| Caliper (internal & External) | 0 mm to 300 mm | 0.006 mm | Calibrated by using caliper checker and gauge block with reference JIS B 7507:2016 |
| Micrometer (external) | 0 mm to 75 mm 75 mm to 150 mm 150 mm to 200 mm | 0.002 mm 0.003 mm 0.004 mm | Calibrated by using gauge block with reference JIS B 7502:2016 |

| | |
|----------------------------------|------------------------------------|
| SITE LOCATION (HQ) | 1. CATEGORY I |
| FIELD(S) OF CALIBRATION : | HEAT & TEMPERATURE, MASS, PRESSURE |

SCOPE OF CALIBRATION : HEAT & TEMPERATURE

| Material / Product Tested | Type Of Test / Properties Measured / Range Of Measurement | Standard Test Methods / Equipment / Techniques | Remarks |
|--|---|--|---|
| Temperature Sensor Pt 100 | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.2 °C 0.4 °C 0.5 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |
| Temperature Sensor Thermocouple | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.2 °C 0.6 °C 0.7 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |
| Temperature Sensor Temperature Gauge | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.3 °C 0.5 °C 0.6 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |

Schedule

Issue date: 12 August 2025
Valid Until: -



NO: SAMM 407

(Issue 2, 12 August 2025 replacement of SAMM 407 dated 14 October 2025)

Page: 6 of 7

| Material / Product Tested | Type Of Test / Properties Measured / Range Of Measurement | Standard Test Methods / Equipment / Techniques | Remarks |
|--|---|--|---|
| Temperature Sensor With Indicator Pt 100 | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.1 °C 0.4 °C 0.5 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |
| Temperature Sensor With Indicator Thermocouple | 0 °C 30 °C to 200 °C 200 °C to 400 °C | 0.1 °C 0.6 °C 0.7 °C | Comparison with Pt100 reference in liquid bath and temperature block calibrator |
| Temperature Measuring Device (by Electrical Simulation) A) Type K | -200 °C to -100 °C -100 °C to 1300 °C | 1.2 °C 0.2 °C | Calibrated by electrical simulation using calibrator and reference table ITS 90 |
| Temperature Measuring Device (by Electrical Simulation) B) Type J | -200 °C to 1200 °C | 0.3 °C | Calibrated by electrical simulation using calibrator and reference table ITS 90 |
| Temperature Measuring Device (by Electrical Simulation) C) Pt 100 | -200 °C to 850 °C | 0.2 °C | Calibrated by electrical simulation using calibrator and reference table ITS 90 |
| Temperature Controlled Enclosure | -40 °C to 250 °C 250 °C to 600 °C 600 °C to 1300 °C | 0.6 °C 1.7 °C 3.2 °C | Calibrated by using temperature recorder with thermocouple |
| Humidity Chamber @ 25 °c | 5 °C to 60 °C 30 %rh to 95 %rh | 0.6 °C 2.2 %rh | Calibrated by using temperature recorder with thermocouple and Data Logger |

SCOPE OF CALIBRATION : MASS

| Material / Product Tested | Type Of Test / Properties Measured / Range Of Measurement | Standard Test Methods / Equipment / Techniques | Remarks |
|---------------------------|--|---|--------------------------------------|
| Balance | Up to 200 g 200 g to 1000 g 1 kg to 5 kg 5 kg to 10 kg 10 kg to 20 kg 20 kg to 50 kg 50 kg to 100 kg 100 kg to 300 kg | 0.7 mg 6 mg 0.03 g 0.07 g 5 g 10 g 0.02 kg 0.05 kg | Calibrated by using standard weights |

NO: SAMM 407(Issue 2, 12 August 2025 replacement
of SAMM 407 dated 14 October 2025)

Page: 7 of 7

SCOPE OF CALIBRATION : PRESSURE

| Material / Product Tested | Type Of Test / Properties Measured / Range Of Measurement | Standard Test Methods / Equipment / Techniques | Remarks |
|--|---|---|-----------------------------------|
| Pressure Measuring Device A) Hydraulic | 0 bar to 300 bar 300 bar to 700 bar | 0.07 bar 0.13 bar | Calibrated by using test gauge |
| Pressure Measuring Device B) Pneumatic | 0 bar to 0.005 bar 0.005 bar to 0.01 bar 0.01 bar 0.03 bar 0.03 bar to 0.05 bar 0 bar to 30 bar | 0.004 mbar 0.006 mbar 0.02 mbar 0.04 mbar 0.005 bar | Calibrated by using test gauge |
| Pressure Measuring Device C) Vacuum | -0.9 bar to 0 bar | 0.002 bar | Calibrated by using test gauge |