

NO: SAMM 669

Page: 1 of 1

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	EDMI Electronics Sdn. Bhd. PLO 226A, Jalan Cyber 1A Kawasan Perindustrian Senai Fasa III 81400 Senai, Johor , 81400, JOHOR MALAYSIA
	
<b>ACCREDITED SINCE :</b>	17 OCTOBER 2014
<b>FIELD(S) OF TESTING:</b>	MECHANICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

<b>CENTRAL LOCATION:</b>	EDMI Electronics Sdn. Bhd. PLO 226A, Jalan Cyber 1A Kawasan Perindustrian Senai Fasa III 81400 Senai, Johor , 81400, Johor
<b>FIELD(S) OF TESTING :</b>	MECHANICAL,

**SCOPE OF TESTING : MECHANICAL**

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
*electronic Gas Meter	Error of indication for flow rate in the range (0.03 to 25) m <sup>3</sup> /h using air medium	In-house method WI-LAB-09 with reference to BS EN 14236:2018 and OIML R137-2012 parts 1 and 2
**electronic Metering Device	Accelerated Life Testing (Temperature and Humidity)	In-house method WI-REL-04 with reference to IEC 62059-31-1 (Edition 1.0 2008-10)
***metal Enclosure (electronic Gas Meter)	Accelerated Life Testing (Cyclic Temperature)	In-house method WI-REL-29 with reference to IEC 60068-2-14:2009 (clause 8)