

# Schedule


Issue date: 24 June 2025  
Valid Until: -



## NO: SAMM 993

(Issue 2, 24 June 2025 replacement  
of SAMM 993 dated 24 June 2025)

Page: 1 of 1

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	Calibration Laboratory, G7 Aerospace Sdn. Bhd. A-2-02 Co Place 1 2270 Jalan Usahawan 2 63000 Cyberjaya, Selangor , 63000, SELANGOR MALAYSIA
	
<b>ACCREDITED SINCE :</b>	24 JUNE 2025
<b>FIELD(S) OF CALIBRATION:</b>	TORQUE

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

**\* The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

<b>CENTRAL LOCATION</b>	Calibration Laboratory, G7 Aerospace Sdn. Bhd. A-2-02 Co Place 1 2270 Jalan Usahawan 2 63000 Cyberjaya, Selangor , 63000, Selangor
<b>FIELD(S) OF CALIBRATION :</b>	TORQUE,

## SCOPE OF CALIBRATION : TORQUE

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
Hand Torque Tools	4 N.m to < 25 N.m	2.0 % of reading	With reference to ISO 6789-2:2017.
	25 N.m to 650 N.m	1.0 % of reading	With reference to ISO 6789-2:2017.

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003309> for the current scope of accreditation