


NO: SAMM 881

Page: 1 of 1

|   |  |
|---|--|
| <b>LABORATORY LOCATION/<br/>CENTRAL OFFICE:</b>                                   | Encord Sdn. Bhd.<br>Lot 2, Jalan Kemajuan Satu 16/17A, Seksyen 16, 40000 Shah Alam,<br>Selangor , 40000,<br>SELANGOR<br>MALAYSIA |
|  |  |
| <b>ACCREDITED SINCE :</b>   | 12 MARCH 2025  |
| <b>FIELD(S) OF TESTING:</b>   | MECHANICAL   |

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

|                              |  |
|------------------------------|--|
| <b>CENTRAL LOCATION:</b>     | Encord Sdn. Bhd.<br>Lot 2, Jalan Kemajuan Satu 16/17A, Seksyen 16, 40000 Shah Alam,<br>Selangor , 40000,<br>Selangor |
| <b>FIELD(S) OF TESTING :</b> | MECHANICAL,  |

**SCOPE OF TESTING : MECHANICAL**

| <b>Material / Product Tested</b>            | <b>Type Of Test / Properties Measured / Range Of Measurement</b> | <b>Standard Test Methods / Equipment / Techniques</b> |
|---|--|---|
| Safety Relief Valve (pressure Up To 60 Bar) | 1. Pre-Pop Pressure Test (As Received)                           | 1. API 576 (Fourth Edition 2017) Clause 6.2.9         |
|   | 2. Seat Tightness Test   | 2. API 527 (Fifth Edition 2020) Section 1, 2, 4 & 5   |
|   | 3. Final Set Pressure Test                                       | 3. API 576 (Fourth Edition 2017) Clause 6.2.9.8       |