


Schedule

Issue date: 19 March 2025
Valid Until: -



NO: SAMM 1042

Page: 1 of 1

LABORATORY LOCATION/ CENTRAL OFFICE:	Supergenics Solution Sdn Bhd Suite A-01-03, The Palace @One City Jalan USJ 25/1A 47650 Subang Jaya, Selangor, 47650, SELANGOR MALAYSIA
	
ACCREDITED SINCE :	19 MARCH 2025
FIELD(S) OF CALIBRATION:	HEAT AND TEMPERATURE

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

*** The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

CENTRAL LOCATION	Supergenics Solution Sdn Bhd Suite A-01-03, The Palace @One City Jalan USJ 25/1A 47650 Subang Jaya, Selangor, 47650, Selangor
FIELD(S) OF CALIBRATION :	HEAT & TEMPERATURE,

SCOPE OF CALIBRATION : HEAT AND TEMPERATURE

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Temperature And Humidity Meter	to 50°C	0.7°C	thermohygrometer in programmable

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003596> for the current scope of accreditation