


# Schedule

Issue date: 19 March 2025  
Valid Until: -



NO: SAMM 1091

Page: 1 of 1

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	Mitsudell Solutions Sdn. Bhd. 18-3-15 & 18-3-17, Gembira Parade Lorong Delima 1, Green Lane 11600 Pulau Pinang , 11600, PULAU PINANG MALAYSIA
	
<b>ACCREDITED SINCE :</b>	19 MARCH 2025
<b>FIELD(S) OF CALIBRATION:</b>	HEAT AND TEMPERATURE

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

**\* The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

<b>CENTRAL LOCATION</b>	Mitsudell Solutions Sdn. Bhd. 18-3-15 & 18-3-17, Gembira Parade Lorong Delima 1, Green Lane 11600 Pulau Pinang , 11600, Pulau Pinang
<b>FIELD(S) OF CALIBRATION :</b>	HEAT & TEMPERATURE,

## SCOPE OF CALIBRATION : HEAT AND TEMPERATURE

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
Melt Flow Indexer A) Barrel - Temperature	190 °C 300 °C	0.8 °C	Indexer Series 4000 & 5000 by comparison method with

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003624> for the current scope of accreditation