


Schedule

Issue date: 24 March 2025
Valid Until: -



NO: SAMM 693

Page: 1 of 15

LABORATORY LOCATION/ CENTRAL OFFICE:	Radio & Frequency (RF) Laboratory, SIRIM Calibration Sdn. Bhd. Bangunan Utama, SIRIM Berhad 483, Mukim 6, Jalan Permatang Pauh, 13500 Permatang Pauh Seberang Prai, Pulau Pinang , 13500, PULAU PINANG MALAYSIA
	
ACCREDITED SINCE :	24 MARCH 2025
FIELD(S) OF CALIBRATION:	ELECTRICAL (DC & LOW FREQUENCY) ELECTRICAL (RF/MICROWAVE (50 ? SYSTEM))

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

*** The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

CENTRAL LOCATION	Radio & Frequency (RF) Laboratory, SIRIM Calibration Sdn. Bhd. Bangunan Utama, SIRIM Berhad 483, Mukim 6, Jalan Permatang Pauh, 13500 Permatang Pauh Seberang Prai, Pulau Pinang , 13500, Pulau Pinang
FIELD(S) OF CALIBRATION :	ELECTRICAL, ELECTRICAL

SCOPE OF CALIBRATION : ELECTRICAL (DC & LOW FREQUENCY)

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Measuring Instrument Dc Voltage	0 to 330 mV 0 to 3.3 V 0 to 33 V 33 to 330 V 330 to 1000 V	(of reading) 7.5 μ V 0.2 mV 0.8 mV 8 mV 74 mV	Generation using calibrator model Fluke 5500A

Schedule

Issue date: 24 March 2025
Valid Until: -



NO: SAMM 693

Page: 2 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Measuring Instrument Dc Resistance	0 to 11 ? 11 to 33 ? 33 to 110 ? 110 to 330 ? 0.33 to 1.1 k ? 1.1 to 3.3 k ? 3.3 to 11 k ? 11 to 33 k ? 33 to 110 k ? 110 to 330 k ? 0.33 to 1.1 M ? 1.1 to 3.3 M ? 3.3 to 11 M ? 11 to 33 M ? 33 to 110 M ? 110 to 330 M ?	(of reading) 7 m ? 7 m ? 7 m ? 8 ? 0.2 ? 1 ? 0.7 ? 0.8 ? 11 ? 10 ? 0.1 k ? 0.2 k ? 0.7 k ? 12 k ? 17 k ? 67 k ?	Generation using calibrator model Fluke 5500A
Measuring Instrument Dc Current	0 to 3.3 mA 0 to 33 mA 0 to 330 mA 0 to 2.2 A 0 to 11 A	(of reading) 0.1 μ A 2.1 μ A 0.1 mA 1.2 mA 7.8 mA	Generation using calibrator model Fluke 5500A

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation

Schedule

Issue date: 24 March 2025
Valid Until: -



NO: SAMM 693

Page: 3 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Measuring Instrument Ac Current	0.03 to 0.33 mA	0.3 μ A	Generation using calibrator model Fluke 5500A
	10 Hz to 20 Hz	0.3 μ A	
	20 Hz to 45 Hz	0.3 μ A	
	45 Hz to 1 kHz	0.3 μ A	
	1 kHz to 5 kHz	0.3 μ A	
	5 kHz to 10 kHz	0.3 μ A	
	0.33 to 3.3 mA	3 μ A	
	10 Hz to 45 Hz	3 μ A	
	45 Hz to 1 kHz	3 μ A	
	1 kHz to 5 kHz	3 μ A	
	5 kHz to 10 kHz	3 μ A	
	3.3 to 33 mA	28 μ A	
	10 Hz to 45 Hz	28 μ A	
	45 Hz to 1 kHz	28 μ A	
	1 kHz to 5 kHz	28 μ A	
	5 kHz to 10 kHz	28 μ A	
	33 to 330 mA	0.4 mA	
	10 Hz to 45 Hz	0.4 mA	
	45 Hz to 1 kHz	0.4 mA	
	1 kHz to 5 kHz	0.4 mA	
	5 kHz to 10 kHz	0.4 mA	
0.33 to 2.2 A	4 mA		
10 Hz to 45 Hz	4 mA		
45 Hz to 1 kHz	7 mA		
1 kHz to 5 kHz	7 mA		
2.2 to 11 A	12 mA		
45 Hz to 65 Hz	13 mA		
65 Hz to 500 Hz	14 mA		
500 Hz to 1 kHz	14 mA		

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/lab/samm-ct/3003752> for the current scope of accreditation

NO: SAMM 693

Page: 4 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Measuring Instrument Ac Voltage	1.0 to 33 mV	3.6 μ V	Generation using calibrator model Fluke 5500A
	10 Hz to 45 Hz	43 μ V	
	45 Hz to 10 kHz	3.6 μ V	
	10 kHz to 20 kHz	5.7 μ V	
	20 kHz to 50 kHz	11 μ V	
	50 kHz to 100 kHz	26 μ V	
	100 kHz to 500 kHz		
	33 to 330 mV	13 μ V	
	10 Hz to 45 Hz	77 μ V	
	45 Hz to 10 kHz	13 μ V	
	10 kHz to 20 kHz	18 μ V	
	20 kHz to 50 kHz	45 μ V	
	50 kHz to 100 kHz	110 μ V	
	100 kHz to 500 kHz		
	0.33 to 3.3 V	0.1 mV	
	10 Hz to 45 Hz	0.1 mV	
	45 Hz to 10 kHz	0.1 mV	
	10 kHz to 20 kHz	0.2 mV	
	20 kHz to 50 kHz	0.2 mV	
	50 kHz to 100 kHz	1.1 mV	
	100 kHz to 500 kHz		
	3.3 to 33 V	1.2 mV	
	10 Hz to 45 Hz	1.1 mV	
	45 Hz to 10 kHz	1.2 mV	
10 kHz to 20 kHz	2.1 mV		
20 kHz to 50 kHz	2.8 mV		
50 kHz to 100 kHz			
33 to 330 V	17 mV		
10 Hz to 45 Hz	14 mV		
45 Hz to 10 kHz	14 mV		
10 kHz to 20 kHz			
330 to 1020 V	39 mV		
45 Hz to 1 kHz	52 mV		
1 kHz to 5 kHz			

Schedule

Issue date: 24 March 2025
Valid Until: -



NO: SAMM 693

Page: 5 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Sourcing/generating Instrument Dc Voltage	100 mV + 100 μ V to + 100 mV - 100 mV to - 100 μ V	(of reading) 1.8 μ V 1.8 μ V	HP3458A
	1 V + 100 mV to 1 V -1 V to -100 mV	9 μ V 9 μ V	
	10 V + 1 V to + 10 V - 10 V to -1 V	0.1 mV 0.1 mV	
	100 V + 10 V to +100 V -100 V to -10 V	1 mV 1 mV	
	1000 V + 100 V to + 1000 V - 1000 V to -100 V	12 mV 12 mV	
Sourcing/generating Instrument Dc Resistance	10 ? 100 ? 1 k ? 10 k ? 100 k ? 1 M ? 10 M ? 100 M ?	0.3 m ? 2.0 m ? 15 m ? 0.1 ? 1.6 ? 23 ? 0.5 k? 13 k?	HP3458A

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation

NO: SAMM 693

Page: 6 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Sourcing/generating Instrument Dc Current	100 μ A +10 μ A to +100 μ A -100 μ A to - 10 μ A 1 mA +100 μ A to +1 mA -1 mA to -100 μ A 10 mA +1 mA to +10 mA -10 mA to -1 mA 100 mA +10 mA to +100 mA -100 mA to -10 mA 1 A +100 mA to + 1 A -1 A to -100 mA	(of reading) 16 nA 16 nA 69 nA 69 nA 0.7 μ A 0.7 μ A 7.8 μ A 7.8 μ A 0.1 mA 0.1 mA	HP 3458A
Sourcing/generating Instrument Ac Voltage	10 mV range 0 mV to 10 mV 100 mV range 10 mV to 100 mV 1 V range 100 mV to 1 V 10 V range 1 V to 10 V 100 V range 10 V to 100 V 1000 V range 100 V to 700 V (See Matrix A)	(See Matrix A)	HP 3458A
High Current Generating Dc Current	1 A to 10 A	0.03 A	HP 6060B / 6050A w/ 60503A module

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation

NO: SAMM 693

Page: 7 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Oscilloscope Vertical Deflection Dc Signal	0 V to \pm 6.6 V (50 Ω Load)	(of reading) 1 mV/V + 0.04 mV	Generating using Oscilloscope Calibrator Fluke 5820A & E8257D
	0 V to \pm 130 V (1 M Ω Load)	0.5 mV/V + 0.04 mV	
Oscilloscope Vertical Deflection Square Wave Signal	\pm 1 mVp-p to \pm 6.6 Vp-p (50 Ω Load)	1.0 mVp-p/Vp-p +0.04 mV	
	\pm 1 mVp-p to 130 Vp-p (1 M Ω Load)	0.5 mVp-p/Vp-p + 0.04 mV	
Oscilloscope Horizontal Deflection Time Markers (50 Ohm Load)	2 ns/div to 20 ms/div	2.5 us/s	
	50 ms/div to 5 s/div	2.2 ms/s	
Oscilloscope Rise Time	\leq 300ps	0.3 ns	
Oscilloscope Bandwidth Amplitude	50 kHz to 600 MHz	17 mVp-p	
	600 MHz to 2100 MHz	23 mVp-p	
Audio Analyzer	Frequency 20 Hz to 150 kHz	0.006 kHz	Fluke 5500A
	Amplitude 1 mV to 300 V	0.11 V	
	Residual Distortion (-99 to 0 dB) 20 Hz to 20 kHz 20 kHz to 100 kHz	1 dB 1 dB	8903B

SCOPE OF CALIBRATION : ELECTRICAL (RF/MICROWAVE (50 Ω SYSTEM))

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Measuring Instrument Dc Voltage	0 to 330 mV	(of reading)	Generation using calibrator model Fluke 5500A
	0 to 3.3 V	7.5 μ V	
	0 to 33 V	0.2 mV	
	33 to 330 V	0.8 mV	
	330 to 1000 V	8 mV 74 mV	

Schedule

Issue date: 24 March 2025
Valid Until: -



NO: SAMM 693

Page: 8 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Measuring Instrument Dc Resistance	0 to 11 ? 11 to 33 ? 33 to 110 ? 110 to 330 ? 0.33 to 1.1 k ? 1.1 to 3.3 k ? 3.3 to 11 k ? 11 to 33 k ? 33 to 110 k ? 110 to 330 k ? 0.33 to 1.1 M ? 1.1 to 3.3 M ? 3.3 to 11 M ? 11 to 33 M ? 33 to 110 M ? 110 to 330 M ?	(of reading) 7 m ? 7 m ? 7 m ? 8 ? 0.2 ? 1 ? 0.7 ? 0.8 ? 11 ? 10 ? 0.1 k ? 0.2 k ? 0.7 k ? 12 k ? 17 k ? 67 k ?	Generation using calibrator model Fluke 5500A
Measuring Instrument Dc Current	0 to 3.3 mA 0 to 33 mA 0 to 330 mA 0 to 2.2 A 0 to 11 A	(of reading) 0.1 μ A 2.1 μ A 0.1 mA 1.2 mA 7.8 mA	Generation using calibrator model Fluke 5500A

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation

NO: SAMM 693

Page: 9 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Measuring Instrument Ac Current	0.03 to 0.33 mA	0.3 μ A	Generation using calibrator model Fluke 5500A
	10 Hz to 20 Hz	0.3 μ A	
	20 Hz to 45 Hz	0.3 μ A	
	45 Hz to 1 kHz	0.3 μ A	
	1 kHz to 5 kHz	0.3 μ A	
	5 kHz to 10 kHz	0.3 μ A	
	0.33 to 3.3 mA	3 μ A	
	10 Hz to 45 Hz	3 μ A	
	45 Hz to 1 kHz	3 μ A	
	1 kHz to 5 kHz	3 μ A	
	5 kHz to 10 kHz	3 μ A	
	3.3 to 33 mA	28 μ A	
	10 Hz to 45 Hz	28 μ A	
	45 Hz to 1 kHz	28 μ A	
	1 kHz to 5 kHz	28 μ A	
	5 kHz to 10 kHz	28 μ A	
	33 to 330 mA	0.4 mA	
	10 Hz to 45 Hz	0.4 mA	
	45 Hz to 1 kHz	0.4 mA	
	1 kHz to 5 kHz	0.4 mA	
	5 kHz to 10 kHz	0.4 mA	
0.33 to 2.2 A	4 mA		
10 Hz to 45 Hz	4 mA		
45 Hz to 1 kHz	7 mA		
1 kHz to 5 kHz	7 mA		
2.2 to 11 A	12 mA		
45 Hz to 65 Hz	13 mA		
65 Hz to 500 Hz	14 mA		
500 Hz to 1 kHz	14 mA		

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation

NO: SAMM 693

Page: 10 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Measuring Instrument Ac Voltage	1.0 to 33 mV	3.6 μ V	Generation using calibrator model Fluke 5500A
	10 Hz to 45 Hz	43 μ V	
	45 Hz to 10 kHz	3.6 μ V	
	10 kHz to 20 kHz	5.7 μ V	
	20 kHz to 50 kHz	11 μ V	
	50 kHz to 100 kHz	26 μ V	
	100 kHz to 500 kHz		
	33 to 330 mV	13 μ V	
	10 Hz to 45 Hz	77 μ V	
	45 Hz to 10 kHz	13 μ V	
	10 kHz to 20 kHz	18 μ V	
	20 kHz to 50 kHz	45 μ V	
	50 kHz to 100 kHz	110 μ V	
	100 kHz to 500 kHz		
	0.33 to 3.3 V	0.1 mV	
	10 Hz to 45 Hz	0.1 mV	
	45 Hz to 10 kHz	0.1 mV	
	10 kHz to 20 kHz	0.2 mV	
	20 kHz to 50 kHz	0.2 mV	
	50 kHz to 100 kHz	1.1 mV	
	100 kHz to 500 kHz		
	3.3 to 33 V	1.2 mV	
	10 Hz to 45 Hz	1.1 mV	
	45 Hz to 10 kHz	1.2 mV	
10 kHz to 20 kHz	2.1 mV		
20 kHz to 50 kHz	2.8 mV		
50 kHz to 100 kHz			
33 to 330 V	17 mV		
10 Hz to 45 Hz	14 mV		
45 Hz to 10 kHz	14 mV		
10 kHz to 20 kHz			
330 to 1020 V	39 mV		
45 Hz to 1 kH	52 mV		
1 kHz to 5 kHz			

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation

Schedule

Issue date: 24 March 2025
Valid Until: -



NO: SAMM 693

Page: 11 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Sourcing/generating Instrument Dc Voltage	100 mV + 100 μ V to + 100 mV - 100 mV to - 100 μ V 1 V + 100 mV to 1 V -1 V to -100 mV 10 V + 1 V to + 10 V - 10 V to -1 V 100 V + 10 V to +100 V -100 V to -10 V 1000 V + 100 V to + 1000 V - 1000 V to -100 V	(of reading) 1.8 μ V 1.8 μ V 9 μ V 9 μ V 0.1 mV 0.1 mV 1 mV 1 mV 12 mV 12 mV	HP3458A
Sourcing/generating Instrument Dc Resistance	10 ? 100 ? 1 k ? 10 k ? 100 k ? 1 M ? 10 M ? 100 M ?	0.3 m ? 2.0 m ? 15 m ? 0.1 ? 1.6 ? 23 ? 0.5 k? 13 k?	HP3458A

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation

NO: SAMM 693

Page: 12 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Sourcing/generating Instrument Dc Current	100 μ A +10 μ A to +100 μ A -100 μ A to - 10 μ A 1 mA +100 μ A to +1 mA -1 mA to -100 μ A 10 mA +1 mA to +10 mA -10 mA to -1 mA 100 mA +10 mA to +100 mA -100 mA to -10 mA 1 A +100 mA to + 1 A -1 A to -100 mA	(of reading) 16 nA 16 nA 69 nA 69 nA 0.7 μ A 0.7 μ A 7.8 μ A 7.8 μ A 0.1 mA 0.1 mA	HP 3458A
Sourcing/generating Instrument Ac Voltage	10 mV range 0 mV to 10 mV 100 mV range 10 mV to 100 mV 1 V range 100 mV to 1 V 10 V range 1 V to 10 V 100 V range 10 V to 100 V 1000 V range 100 V to 700 V (See Matrix A)	(See Matrix A)	HP 3458A
High Current Generating Dc Current	1 A to 10 A	0.03 A	HP 6060B / 6050A w/ 60503A module

Schedule

Issue date: 24 March 2025
Valid Until: -



NO: SAMM 693

Page: 13 of 15

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Oscilloscope Vertical Deflection Dc Signal	0 V to \pm 6.6 V (50 Ω Load)	(of reading) 1 mV/V + 0.04 mV	Generating using Oscilloscope Calibrator Fluke 5820A & E8257D
	0 V to \pm 130 V (1 M Ω Load)	0.5 mV/V + 0.04 mV	
Oscilloscope Vertical Deflection Square Wave Signal	\pm 1 mVp-p to \pm 6.6 Vp-p (50 Ω Load)	1.0 mVp-p/Vp-p +0.04 mV	
	\pm 1 mVp-p to 130 Vp-p (1 M Ω Load)	0.5 mVp-p/Vp-p + 0.04 mV	
Oscilloscope Horizontal Deflection Time Markers (50 Ohm Load)	2 ns/div to 20 ms/div	2.5 us/s	
	50 ms/div to 5 s/div	2.2 ms/s	
Oscilloscope Rise Time	\leq 300ps	0.3 ns	
Oscilloscope Bandwidth Amplitude	50 kHz to 600 MHz	17 mVp-p	
	600 MHz to 2100 MHz	23 mVp-p	
Audio Analyzer	Frequency 20 Hz to 150 kHz	0.006 kHz	Fluke 5500A
	Amplitude 1 mV to 300 V	0.11 V	
	Residual Distortion (-99 to 0 dB) 20 Hz to 20 kHz 20 kHz to 100 kHz	1 dB 1 dB	8903B

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation

NO: SAMM 693

Page: 14 of 15

SCOPE OF CALIBRATION : ELECTRICAL (DC & LOW FREQUENCY)

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Rf Measuring Equipment Frequency Counter	Time Base	(of reading)	Fluke 910R. (GPS-disciplined.)
	10 MHz	0.1 mHz	
	Frequency Range	0.8 mHz	HP 3325A/B HP
	100 Hz to 60 MHz	0.8 mHz	8642A/B HP 83620A
	0.1 MHz to 2.1 GHz	0.8 Hz	E8257D (GPS-disciplined.)
	10 MHz to 20 GHz	0.8 Hz	
Modulation Analyzer	250 kHz to 40 GHz		
	Frequency	0.8 Hz	HP8642A/B
	150kHz to 1.3GHz		
	Power Level	0.06 dBm	HP 11683A
	-25 dBm to 20 dBm (3 μ W to 100 mW)		
	FM, Flatness	0.1 kHz	HP 11715A
	20 Hz to 100 kHz rates	0.5 kHz	
	100 kHz to 200 kHz rates		
	AM, Flatness	0.05 kHz	HP 11715A
	50 Hz to 50 kHz 20 Hz to 100 kHz	0.25 kHz	
Rf Power Meter	Distortion	1 dB	HP 8903B
	Audio Distortion -70 dB minimum		
Network Analyzer	Power Level	0.06 dBm	HP 11683A
	-25 dBm to 20 dBm (3 μ W to 100 mW)		
Network Analyzer	Frequency	(of reading)	E4448A
	9 kHz to 26.5 GHz	0.18 Hz	
	Power Level		
	0 dBm to +20 dBm	0.31 dB	
	9 kHz to 2 GHz	0.57 dB	
	>2 GHz to 8 GHz	0.60 dB	
	>8 GHz to 13.5 GHz	0.67 dB	
	>13.5 GHz to 26.5 GHz		
		0.30 dB	
	0 dBm to -30 dBm	0.57 dB	
	9 kHz to 2 GHz	0.60 dB	
	>2 GHz to 8 GHz	0.64 dB	
	>8 GHz to 13.5 GHz		
	>13.5 GHz to 26.5 GHz		

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation

Schedule

Issue date: 24 March 2025
Valid Until: -



NO: SAMM 693

Page: 15 of 15

Rf Power Sensor 100 Khz To 4.2 Ghz	Power Level	(of reading)	Calibration Factor in a 50 ? system, Type-N(m) 50 ? Calibrated at 1 mW input power, referenced to 50 MHz at 1 mW
	-30 dBm to +20 dBm		
	100 kHz to 300 kHz	(%)	
	300 kHz to 1 MHz	1	
	3 MHz to 30 MHz	1	
	100 MHz to 3.5 GHz	1	
	3.7 GHz to 4.2 GHz	1	
		1	
Rf Power Sensor 10 Mhz To 18 Ghz	Power Level		
	-30 dBm to +20 dBm		
	10 MHz to 300 MHz	2	
	300 MHz to 1.5 GHz	2	
	1.5 GHz to 8 GHz	2	
	8 GHz to 13 GHz	2	
	13 GHz to 18 GHz		
Rf Power Sensor 100 Khz To 4.0 Ghz	Power Level		
	-10 dBm to +35 dBm		
	100 kHz to 500 kHz	1	
	500 kHz to 10 MHz	1	
	10 MHz to 50 MHz	1	
	50 MHz to 2 GHz	1	
Spectrum Analyzer	2 GHz to 4 GHz		
	Frequency Range	0.16 Hz	HP 83620A HP
	10 MHz to 20 GHz	0.18 Hz	83630A E8257D
	10 MHz to 26.5 GHz	0.18 Hz	(GPS-disciplined)
	250 kHz to 40 GHz		
	Resolution Bandwidth	1 MHz	
1 kHz to 3 MHz			
Reference Level	1 dB	E8257D	
Freq.: 250 kHz to 40 GHz			
	-90 dBm to 0 dBm		
Rf Generating Equipment Signal Generator	150 kHz to 1300 MHz	3.0 x 10 ⁻⁵ Hz	HP 8902A w/ 11722A
	1 Hz to 12 GHz	3.0 x 10 ⁻⁵ Hz	HP 53132A E4448A
	Frequency	9 kHz to 50 GHz	3.0 x 10 ⁻⁵ Hz
Rf Generating Equipment Amplitude Modulation (am) depth: 10 % To 90 % rate: 50hz To 10 Khz Rate: 50hz To 100khz	150 kHz to 10 MHz	4 % of reading	HP 8902A w/ 11722A
	10 MHz to 1300 MHz	4 % of reading	

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003752> for the current scope of accreditation