


NO: SAMM 1115

Page: 1 of 2

LABORATORY LOCATION/ CENTRAL OFFICE:	J.I.L. Logistics & Engineering (S) Sdn. Bhd. Lot 1185 & 1269, Block 12 (New Lot 1423), No.93-B, Jalan Setia Raja, Muara Tabuan Light Industrial Park, 93350 Kuching, Sarawak , 93350, SARAWAK MALAYSIA
	
ACCREDITED SINCE :	24 MARCH 2025
FIELD(S) OF TESTING:	MECHANICAL (VALVE TESTING)
FIELD(S) OF CALIBRATION:	PRESSURE

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

*** The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

CENTRAL LOCATION:	J.I.L. Logistics & Engineering (S) Sdn. Bhd. Lot 1185 & 1269, Block 12 (New Lot 1423), No.93-B, Jalan Setia Raja, Muara Tabuan Light Industrial Park, 93350 Kuching, Sarawak , 93350, Sarawak
FIELD(S) OF TESTING :	MECHANICAL,

SCOPE OF TESTING : MECHANICAL (VALVE TESTING)

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Pressure Relief Valves	-œAs Received-• Pop Pressure Test	API 576 Clause 6.2.9: 2017

CENTRAL LOCATION	J.I.L. Logistics & Engineering (S) Sdn. Bhd. Lot 1185 & 1269, Block 12 (New Lot 1423), No.93-B, Jalan Setia Raja, Muara Tabuan Light Industrial Park, 93350 Kuching, Sarawak , 93350, Sarawak
-------------------------	---

Schedule

Issue date: 24 March 2025
Valid Until: -



NO: SAMM 1115

Page: 2 of 2

FIELD(S) OF CALIBRATION :	PRESSURE,
----------------------------------	-----------

SCOPE OF CALIBRATION : PRESSURE

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Hydraulic Pressure Measuring Devices	0 bar to 70 bar	0.55 bar	(04/2019), DKD-R6- 1 (Ed. 03/2014), BS EN 837-1:1998 and BS
Pneumatic Pressure Measuring Devices	0 bar to 70 bar	0.42 bar	Comparison against standard test gauges with general reference to Euramet cg 17 v4.0

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003785> for the current scope of accreditation