Schedule

Issue date: 25 March 2025

Valid Until: -



NO: SAMM 1127

Page: 1 of 2

LABORATORY LOCATION/ CENTRAL OFFICE:	Laboratory, Camfil Malaysia Sdn. Bhd. Plot 9A & 9B, Lorong Bemban 1 Bemban Industrial Estate 31000 Batu Gajah, Perak, 31000, PERAK MALAYSIA
ACCREDITED SINCE :	25 MARCH 2025
FIELD(S) OF TESTING:	CHEMICAL MECHANICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

CENTRAL LOCATION:	Laboratory, Camfil Malaysia Sdn. Bhd.
	Plot 9A & 9B, Lorong Bemban 1 Bemban Industrial Estate 31000
	Batu Gajah, Perak , 31000,
	Perak
FIELD(S) OF TESTING	CHEMICAL MECHANICAL
FIELD(S) OF TESTING:	CHEMICAL, MECHANICAL

SCOPE OF TESTING: CHEMICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Others Activated Carbon	Butane Activity (%)	ASTM D5742-16
	Moisture Content (%)	ASTM D2867-09
	Apparent Density (g/ml))	ASTM D2854-09

SCOPE OF TESTING: MECHANICAL

Schedule

Issue date: 25 March 2025

Valid Until: -



NO: SAMM 1127

Page: 2 of 2

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Air Filter	Classification system based upon particulate matter efficiency (ePM)	ISO 16890-1:2016 (Clause 7.3)
	Pressure Drop (Pa)	ISO 16890-2:2022
	Filter Efficiency (%) (DEHS or KCI)	ISO 16890-2:2022
	Discharged Filter Efficiency (%) (DEHS or KCI)	ISO 16890-4:2022
	Dust Holding Capacity (g)	ISO 16890-3:2016