


Schedule

Issue date: 26 March 2025
Valid Until: -



NO: SAMM 324

Page: 1 of 2

LABORATORY LOCATION/ CENTRAL OFFICE:	I.S. Calibrations Sdn Bhd 12, Persiaran Jelapang Maju 5 Kawasan Perindustrian Ringan Jelapang Maju 30200 Ipoh, Perak , 30200, PERAK MALAYSIA
	
ACCREDITED SINCE :	05 APRIL 2006
FIELD(S) OF CALIBRATION:	MASS (MASS AND MASS RELATED QUANTITIES)

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

*** The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

CENTRAL LOCATION	I.S. Calibrations Sdn Bhd 12, Persiaran Jelapang Maju 5 Kawasan Perindustrian Ringan Jelapang Maju 30200 Ipoh, Perak , 30200, Perak
FIELD(S) OF CALIBRATION :	MASS,

SCOPE OF CALIBRATION : MASS (MASS AND MASS RELATED QUANTITIES)

Schedule

Issue date: 26 March 2025
Valid Until: -



NO: SAMM 324

Page: 2 of 2

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Weighing Balances And Scales	Up to 300g 300g to 600g 600g to 1kg 1kg to 3kg 3kg to 5kg 5kg to 30kg 30kg to 300kg	0.01g 0.02g 0.008g 0.012g 0.580g 0.004kg 0.007kg	1. The calibration procedure covers the repeatability, corner tests and departure from nominal value according to OIML R76-1 (2006) 2. The CMC is estimated from the contribution of the reference standard weight (F2 and M2).

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3003924> for the current scope of accreditation