Schedule

Issue date: 26 March 2025
Valid Until: -

NO: SAMM 350

LABORATORY LOCATION/ CENTRAL OFFICE:	Krizik (M) Sdn Bhd No. 5, Jalan Teras 2 Taman Industri Selesa Jaya 43300 Seri Kembangan, Selangor , 43300, SELANGOR MALAYSIA
ACCREDITED SINCE :	26 MARCH 2025
FIELD(S) OF CALIBRATION:	ELECTRICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

*** The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

CENTRAL LOCATION	Krizik (M) Sdn Bhd No. 5, Jalan Teras 2 Taman Industri Selesa Jaya 43300 Seri Kembangan, Selangor , 43300, Selangor
FIELD(S) OF CALIBRATION :	ELECTRICAL,

SCOPE OF CALIBRATION : ELECTRICAL

Instrument Calibrated/Measurement Parameter	**Range**	**Calibration and Measurement Capability Expressed as an Uncertainty (±)***	**Remarks**
A) 3 Phase	Current: 0.005 A to 100 A	None	comparison with
	Power factor: 0, 0.5, 0.866, 1	None	reference
	Power factor: 0, 0.5, 0.866, 1	None	standard energy

Issue date: 26 March 2025
Valid Until: -

NO: SAMM 350

Instrument Calibrated/Measurement Parameter	**Range**	**Calibration and Measurement Capability Expressed as an Uncertainty (±)***	**Remarks**
	Power factor: 0, 0.5, 0.866, 1	None	measurement
Electronic Meter	Voltage: 60 V to 276 V (Vp-n)	0.04%	Direct
	a) Test of starting and no-load	a) IEC 62053-21 clause 8.3	