

NO: SAMM 386

Page: 1 of 1

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	CLMO Technology Sdn. Bhd. No.10 Jalan Dagang SB 4/2 Taman Sg. Besi Indah 43300 Seri Kembangan, Selangor , 43300, SELANGOR MALAYSIA
	
<b>ACCREDITED SINCE :</b>	14 FEBRUARY 2008
<b>FIELD(S) OF CALIBRATION:</b>	TEMPERATURE

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

**\* The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

<b>CENTRAL LOCATION</b>	CLMO Technology Sdn. Bhd. No.10 Jalan Dagang SB 4/2 Taman Sg. Besi Indah 43300 Seri Kembangan, Selangor , 43300, Selangor
<b>FIELD(S) OF CALIBRATION :</b>	HEAT & TEMPERATURE,

**SCOPE OF CALIBRATION : TEMPERATURE**

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
Chamber	5 %RH to 98 %RH	4 %RH	JTM K 09
Temperature & Humidity	-80 °C to 300 °C	None	Based on JTM K 07 &