

Schedule

Issue date: 06 April 2025
Valid Until: -



NO: SAMM 549

Page: 1 of 2

LABORATORY LOCATION/ CENTRAL OFFICE:	CLMO Technology (Penang) Sdn. Bhd. No. 6, Lintang Bayan Lepas 2 Taman Perindustrian Bayan Lepas Fasa 4 11900 Bayan Lepas, Pulau Pinang , 11900, PULAU PINANG MALAYSIA
	
ACCREDITED SINCE :	06 APRIL 2025
FIELD(S) OF TESTING:	MECHANICAL (ENVIRONMENTAL TESTING)

<p>This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).</p> <p>This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).</p>	
CENTRAL LOCATION:	CLMO Technology (Penang) Sdn. Bhd. No. 6, Lintang Bayan Lepas 2 Taman Perindustrian Bayan Lepas Fasa 4 11900 Bayan Lepas, Pulau Pinang , 11900, Pulau Pinang
FIELD(S) OF TESTING :	MECHANICAL,

SCOPE OF TESTING : MECHANICAL (ENVIRONMENTAL TESTING)

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Semiconductor, Printed Circuit Board Assembly, Finished Electrical And Electronic Product	Damp heat cyclic 10 °C to 85 °C 20 %RH to 98 %RH	IEC 60068-2-30:2008
	Dry heat 30 °C to 200 °C	MS IEC 60068-2-2:2019 (Test Bb)
	Damp heat, steady state 10 °C to 85 °C 20 %RH to 98 %RH	MS IEC 60068-2-78:2012
	Change of temperature Max temperature = 150 °C Min temperature = -65 °C	IEC 60068-2-14:2009 (Test Na and Test Nb)

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/ceb/samm-ct/300407Z> for the current scope of accreditation

Schedule

Issue date: 06 April 2025
Valid Until: -



NO: SAMM 549

Page: 2 of 2

SITE LOCATION (HQ)	1.
FIELD(S) OF CALIBRATION :	HEAT & TEMPERATURE

SCOPE OF CALIBRATION : HEAT & TEMPERATURE

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques	Remarks
Temperature & Humidity Chamber	-80 °C to 300 °C 5 %RH to 98 %RH	0.7 °C 4 %RH	Based on JTM K07:2007 and JTM K09:2009

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/300407Z> for the current scope of accreditation