

Schedule

Issue date: 06 April 2025
Valid Until: -



NO: SAMM 759

Page: 1 of 23

LABORATORY LOCATION/ CENTRAL OFFICE:	Rohde & Schwarz Malaysia Sdn. Bhd. PAT Square, Jalan Pelukis U1/46, Temasya Industrial Park, 40150 Shah Alam, Selangor , 40150, SELANGOR MALAYSIA
	
ACCREDITED SINCE :	06 APRIL 2025
FIELD(S) OF TESTING:	ELECTRICAL ELECTRICAL (RADIO FREQUENCY)

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

CENTRAL LOCATION:	Rohde & Schwarz Malaysia Sdn. Bhd. PAT Square, Jalan Pelukis U1/46, Temasya Industrial Park, 40150 Shah Alam, Selangor , 40150, Selangor
FIELD(S) OF TESTING :	ELECTRICAL, ELECTRICAL

SCOPE OF TESTING : ELECTRICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Ac Current	220 to 2.2A (see Matrix B)	(See Matrix B)
	150 A to 1000 A	None
	1000 Hz:	None
	00to10A 10Ato 100A	1.3 mA/A
	5000 Hz:	None
	00to10A	1.3 mA/A
	10Ato 100A	None
	10000 Hz:	None
	00to10A	1.3 mA/A
	10Ato 100A	None
	See Matrix F	None

Schedule

Issue date: 06 April 2025

Valid Until: -



NO: SAMM 759

Page: 2 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	0 to 33 mA 0 to 330 mA 0to22A 0to11A	None
	0.03 to 0.33 mA	None
	10 Hz to 20 Hz	0.3 pA
	20 Hz to 45 Hz	0.3 pA
	45 Hz to 1 kHz	0.3 pA
	1 kHz to 5 kHz 5 kHz to 10 kHz	0.3 pA 0.3 pA
	0.33 to 3.3 mA	None
	10 Hz to 45 Hz	3 yA
	45 Hz to 1 kHz	3 yA
	1 kHz to 5 kHz	3 yA
	5 kHz to 10 kHz	3 yA
	3.3 to 33 mA	None
	10 Hz to 45 Hz	28
	45 Hz to 1 kHz	28
	1 kHz to 5 kHz	28
	5 kHz to 10 kHz	28
	33 to 330 mA	None
	10 Hz to 45 Hz	0.4mA
	45 Hz to 1 kHz	0.4mA
	1 kHz to 5 kHz	0.4mA
	5 kHz to 10 kHz	0.4mA
	0.33 to 2.2A	None
	10 Hz to 45 Hz	4mA
	45 Hz to 1 kHz	4mA
	1 kHz to 5 kHz	7mA
	2.2to11A	None
	45 Hz to 65 Hz	None
	65 Hz to 500 Hz	13 mA
	500 Hz to 1 KHz	14mA
	10 Hz to 20 Hz 20 Hz to 45 Hz	None
	45 Hz to 1 kHz	None
	1 kHz to 5 kHz	None
	5 kHz to 10 kHz	5 yA
	0.33 to 3.3 mA	None
	10 Hz to 45 Hz	5 yA
	45 Hz to 1 kHz	5 yA
	1 kHz to 5 kHz	None
	5 kHz to 10 kHz	23
	3.3 to 33 mA	None
	10 Hz to 45 Hz	49 pA
	45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	46 pA 83 0.2mA
	33 to 330 mA	None
	10 Hz to 45 Hz	0.6 mA
	45 Hz to 1 kHz	0.5mA
	1 kHz to 5 kHz	0.9mA
	5 kHz to 10 kHz	2.3mA

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3004176> for the current scope of accreditation

NO: SAMM 759

Page: 3 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	0.33 to 2.2A	None
	10 Hz to 45 Hz	6mA
	45 Hz to 1 kHz	5mA
	1 kHz to 5 kHz	20 mA
	2.2to11A	None
	45 Hz to 65 Hz	None
	65 Hz to 500 Hz	20 mA
	500 Hz to 1	45 mA
	33 UA to 10A	See Matrix B
	290 to See Matrix B	See Matrix B
	0.1 20A See Matrix D	See Matrix D
	0.1 UA to 20A See Matrix H	See Matrix H
	See Matrix B	See Matrix B
	See Matrix D	See Matrix D
	100 yA	None
	10 Hz to 1 kHz 1 kHz to 5 kHz	0.0097 pA 0.016 pA
	2.2Ato11A 45 Hz to 65 Hz	1.00mA/A
	65 Hz to 500 Hz	None
	500 Hz to 1 kHz	None
	11 A to 550 A (current coil) 45 Hz to 65 Hz 65 Hz to 440 Hz	3.4mA/A
	100A	None
	1mAto3A	See Matrix D
	1mAto3A	None
	3Ato 100A	None
	45 Hz to 65 Hz	2.4mA/A
	45 Hz to 65 Hz	None
	50 to3A	None
	3Ato 100A	None
	45 Hz to 65 Hz	2.5mA/A
	45 Hz to 65 Hz	None
	45 Hz to 65 Hz	None
Ac Voltage	2.2 mV to 220 V (See Matrix A)	(See Matrix A)
	220 V to 1100 V	None
	15 Hz to 50 Hz	0.30 mV/V + 16 mV
	220 V to 1100 V	None
	50 Hz to 1 kHz	77 + 3.2 mV
	See Matrix E	None
	0.5 kV to 10 kV	12 mV/V + 13 mV
	1.0 to 33 mV	None
	10 Hz to 45 Hz	3.6
	45 Hz to 10 kHz	43
	10 kHz to 20	3.6 UV
	20 kHz to 50 kHz	5.7
	50 kHz to 100 kHz	11
	100 kHz to 500 kHz	26
	33 to 330 mV	None
	10 Hz to 45 Hz	13

NO: SAMM 759

Page: 4 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	45 Hz to 10 kHz	77 uv
	10 kHz to 20	13
	20 kHz to 50 kHz	18
	50 kHz to 100 kHz 100 kHz to 500 kHz	45 110
	0.33 to 3.3 V	None
	10 Hz to 45 Hz	0.1 mV
	45 Hz to 10 kHz	0.1 mV
	10 kHz to 20 kHz	0.1 mV
	20 kHz to 50 kHz	0.2 mV
	50 kHz to 100 kHz	0.2 mV
	100 kHz to 500 kHz	1.1 mV
	3.3 to 33. V	None
	10 Hz to 45 Hz	1.2mV
	45 Hz to 10 kHz	None
	10 kHz to 20 kHz	1.2mV
	20 kHz to 50 kHz	2.1mV
	50 kHz to 100 kHz	2.8 mV
	33 to 330 V	None
	10 Hz to 45 Hz	None
	45 Hz to 10 kHz	14 mV
	10 kHz to 20 kHz	14 mV
	330 to 1020 V	None
	45 Hz to 1 kHz	39 mV
	1 kHz to 5 kHz	52 mV
	-1 Ato -100 mA	None
	10 mV rang	None
	0 mV to 10 mV	None
	100 mV rang	None
	10 mV to 100 mV	None
	1V rang	None
	100 mV to 1 V	None
	10 V range	(See Matrix A)
	1Vto10V	None
	100 V rang	None
	10 V to 100 V	None
	1000 V rang	None
	100 V to 700 V	None
	(See Matrix A)	None
	10 Hz to 45 Hz	0.14 mV
	45 Hz to 10 kHz	85
	10 kHz to 20 kHz 20 kHz to 50 kHz	91 0.11 mV
	50 kHz to 100 kHz	0.16 mV
	100 kHz to 500 kHz	0.41 mV
	33 to 330 mV	None
	10 Hz to 45 Hz	0.9 mV
	45 Hz to 10 kHz	0.2 mV

Schedule

Issue date: 06 April 2025

Valid Until: -



NO: SAMM 759

Page: 5 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	10 kHz to 20 kHz	None
	20 kHz to 50 kHz	0.6 mV
	50 kHz to 100 kHz	1.0 mV
	100 kHz to 500 kHz	2.8 mV
	0.33 to 3.3 V	None
	10 Hz to 45 Hz	5.9 mV
	45 Hz to 10 kHz	1.2mV
	10 kHz to 20 kHz	3.1 mV
	20 kHz to 50 kHz	5.6 mV
	50 kHz to 100 kHz	11 mV
	100 kHz to 500 kHz	22 mV
	3.3 to 33 V	None
	10 Hz to 45 Hz	59 mV
	45 Hz to 10 kHz	16 mV
	10 kHz to 20 kHz	33 mV
	20 kHz to 50 kHz	76 mV
	50 kHz to 100 kHz	110 mV
	33 to 330 V	None
	10 Hz to 45 Hz	0.2
	45 Hz to 10 kHz	0.3
	10 kHz to 20 kHz	0.3
	330 to 1020 V	None
	45 Hz to 1 kHz	0.7
	1 kHz to 5 kHz	2.3
	- 1000 V to -100 V	None
	10 mV rang	None
	0 mV to 10 mV	None
	100 mV range	None
	10 mV to 100 mV	None
	1V rang	None
	100 mV to1V	None
	10 V rang	(See Matrix B)
	1Vto10V	None
	100 V rang	None
	10 V to 100 V	None
	1000 V rang	None
	100 V to 700 V	None
	(See Matrix B)	None
	33 mV to 750 V	See Matrix A
	33 mV to 750 V	None
	33 mV to 1020 V See Matrix A	See Matrix A
	100 to 1050 V See Matrix C	See Matrix C
	33 mV to 1020 V See Matrix E	See Matrix E
	100 to 1050 V See Matrix G	See Matrix G
	See Matrix A	See Matrix A
	See Matrix C	See Matrix C
	10 kHz to 20 kHz	1.0mV/V
	330 V to 1020 V	None

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3004176> for the current scope of accreditation

Schedule

Issue date: 06 April 2025
Valid Until: -



NO: SAMM 759

Page: 6 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	45 Hz to 1 kHz	0.67
	1 kHz to 5 kHz	2.3mV/V
	5 kHz to 10 kHz	2.3mV/V
	1 mV to 750 V	See Matrix C
	750 V to 1000 V 45 Hz to 1 kHz 1 kHz to 10 kHz	9.2mV/V 9.2mV/V
	45 Hz to 65 Hz	35
	45 Hz to 65 Hz	None
	45 Hz to 65 Hz	None
	1 mA to 10 mA	0.90 mA/A
	10 mA to 100 mA	0.77
	0.045 kHz to 1 kHz	9.2mV/V
	1 kHz to 10 kHz	9.2mV/V
	to 40 kV 45 Hz to 65 Hz	62 mV/V
	to 40 kV 45 Hz to 65 Hz	None
Dc Current	0 UA to 220 pA	0.94 nA/A + 7.8 nA
	220 UA to 2.2 mA	8.6 nA/A + 24 nA
	2.2 mA to 22 mA	87 + 0.21
	22 mA to 220 mA	1.5 + 3.4
	220 mA to 2.2A	18 + 48 LA
	10 pA	0.2
	10 to 100 pA	37 +
	100 pA to 1mA	30 + 6 nA
	1 mA to 10 mA	30 + 52 nA
	10 mA to 100 mA	47 + 0.6
	100 mAto1A	0.1 + 5 pA
	0 to 3.3 mA	0.1 pA
	0 to 33 mA 0 to 330 mA 0to22A 0to11A	2.1 0.1mA 1.2mA 7.8mA
	100 pA	None
	+10 to +100 pA	16 nA
	-100 pA 10 pA	16 nA
	+100 to +1 mA	69 nA
	-1 mA to -100 pA	69 nA
	10 mA	None
	+1 +10 mA	0.7
	-10 mA to -1 mA	0.7 UA
	100 mA	None
	+10 mA to +100 mA	None
	-100 mA to -10 mA	None
	1A	None
	+100 mAto+1A	0.1 mA
	-1 Ato -100 mA	0.1 mA
	110 to 330 MQ	(of reading)
	0 to 3.3 mA mA 0 to 330 mA 0to22A	0.5 pA 0.1mA 1.4mA
	0 to 3.3 mA mA 0 to 330 mA 0to22A	11mA

NO: SAMM 759

Page: 7 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	+10 to +100 pA -100 pA 10 pA	16 nA 16 nA
	1mA	None
	+100 A to +1 mA	74nA
	mA to -100 pA	74nA
	10 mA	None
	+1 mA to +10 mA	0.7 pA
	-10 mA to -1 mA	0.7 pA
	100 mA	None
	+10mA to +100 mA	8 yA
	-100 mA to -10 mA	8 yA
	1A	None
	+100 mMAto+1A	0.2mA
	-1 Ato -100 mA	0.2mA
	0 to 330 pA	170 + 23 nA
	330 UA to 3.3 mA	113 + 57 pA
	3.3 mA to 33 mA	113 + 0.29 PA
	33 mA to 330 mA	113 PA/A + 2.8 LA
	330 mA	227 + 45 pA
	1.1Ato3A	429 + 48 LA
	3Ato10A	567 + 562 PA
	3Ato10A	None
	330 A + (0 pA to 330 A)	+ 0.02
	3.3 mA mA to 3.3 mA)	0.12 + 0.05
	33 mA + (0 mA to 33 mA)	0.12 mA/A PA
	330 mA + (0 mA to 330 mA)	0.12
	1.1A +(0Ato	0.23 mA/A + 47 PA
	3A +(1.1Ato3A)	0.46 mAVA + 0.11 mA
	11A	0.58 mA/A + 0.58 mA
	20.5A +(11A to 20.5 A)	1.2 mA/A + 0.87 mA
	200 (0.1 A to 199.99	13 + 0.6 nA
	2mA (0.2 mA to 199.99 mA)	13 + 0.006 pA
	20 mA (2 mA to 19.999 mA)	16 + 0.06
	200 mA (20 mA to 199.99 mA)	55 + 0.98
	2A (0.2 A to 1.9999 A)	0.21 mA/A + 18
	20A (2 A to 19.999 pA)	0.46 mAVA + 0.46 mA
	330 pA + (0 pA to 330 pA)	0.17 mMA/A + 0.02 pA
	3.3 mA mA to 3.3 mA)	0.12 + 0.05 pA
	33 mA + (0 mA to 33 mA)	0.12 0.3 pA
	330 mA + (0 mA to 330 mA)	0.12
	1.1A +(0Ato	0.23 mA/A + 47 pA
	3A	0.46 + 0.11 mA
	11A	0.58 mA/A + 0.58 mA
	20.5A +(11A to 20.5A)	+ 0.87 mA
	200 pA (0.1 pA to 199.99 yA)	13 nA
	2mA (0.2 mA to 199.99 mA)	13 + 0.006
	(2 mA to 19.999 mA)	16 + 0.06 pA
	200 mA (20 mA to 199.99 mA)	55 + 0.98
	2A (0.2 A to 1.9999 A)	0.21 18 pA

NO: SAMM 759

Page: 8 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	20A (2 A to 19.999 pA)	0.46 mA/A + 0.46 mA
	0 to 100 pA 0.1 mA to 1mA	0.058% of 0.035% of rdg + 0.0004 mA
	50 to 500 pA	0.06% of rdg + 0.21
	0 to 100 pV 0 to 0 to 10 mV 0 to 100 mV 0 to 1V 0 to 10 V 0 to 100 V 0 to 1000 V	None
	0 to 100 pA 0 to 1mA 0 to 10 mA 0 to 100 mA 0 to 1A	0.013 pA 0.00013 mA 0.0016 mA 0.028 mA 0.000028 A
	0 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 0.330 A 0.330 A 2.2 A to 11A	0.17 0.13 0.15 mA/A 0.37 mA/A 0.71 mA/A
	0 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 0.330 A 0.330 A 2.2 A to 11A	None
	11 A to 550 A (current coil)	None
	11 A to 550 A (current coil)	None
	0.029 mA to 330 mA	See Matrix B
	0.33 A to 2.2A	None
	10 Hz to 45 Hz	2.5 mA/A
	45 Hz to 1 kHz 1 kHz to 5 kHz	None
	100 mA to 1A 1A to 3A	1.3 mA/A
	100 mA to 1A 1A to 3A	None
	100A	None
	100A	None
	0 mA to 24 mA	0.95 mA/A
	0 mA to 24 mA	None
	0 mA to 24 mA	12 mQ/Q
	10 Q to 100 Q	8.8 mQ/Q
	100 mA to 1A 1A to 3A	1.3 mA/A 1.6 mA/A
	100 mA to 1A 1A to 3A	None
	100A	2.5 mA/A
	100A	None
	100A	None
	100A	None
	50 to 3A	See Matrix F
Dc Voltage	0 V to 220 mV	4.2
	220 mV to 2.2 V	6.1 + 1.1
	2.2 V to 11V	5.1 + 2.0
	11V 22 V to 220 V 220 V to 1100 V	5.0 uV/V + 3.2 7.4 + 28 8.9 + 0.31 mV
	100 mV	None
	+ 100 to + 100 mV	1.8
	-100 mV to - 100	1.8
	iV	None
	+100 mV to 1V	None
	-1 V to -100 mV	None
	10 V	None

NO: SAMM 759

Page: 9 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	+1Vto+10V	0.1 mV
	-10Vto-1V	0.1 mV
	100 V	None
	+10V to +100 V	None
	-100 V to -10 V	1mV/V
	1000 V	None
	+ 100 V to + 1000 V	12 mV
	- 1000 V to -100 V	12 mV
	0 to 330 mV 0to3.3V	30 0.2 mV
	0 to	2.2mV
	33 to 330 V	24 mV
	330 to 1000 V	110 mV
	100 mV	None
	+ 100 to + 100 mV	None
	-100	None
	1V	None
	+100 mV to1V	0.12 mV
	-1 V to -100 mV	0.12 mV
	10V	None
	+1Vto+10V	None
	-10V to-1V	None
	100 V	None
	+10V to +100 V	None
	-100 V to -10 V	None
	1000 V	None
	+ 100 V to + 1000 V	17 mV
	- 1000 V to -100 V	17 mV
	0 to 100 mV	0.004%of rdg+0.004mV
	0 to 50 mV	0.05%of rdg+0.005mV
	30 V to 329 V	0.066 mv /V
	329 V to 1020 V	0.066
	1 mV to 33 V	See Matrix A
	1 mV to 33 V	None
	33 V to 330 V	None
	45 Hz to 1 kHz	0.6 mV/V
	1 kHz to 10 kHz	0.92 mV/V
	1 mV to 100 mV 100 mV to 1V 1Vto10V 10 V to 100 V 100 V to 1000 V	0.098 mV /V 0.054 0.046 0.059 0.064 mV /V
	1 kV to 30 kV	23 mV/V
	1 kV to 30 kV	None
	1 kV to 30 kV	None
	1 kV to 30 kV	None
	0V to	0.58
	1 mV to 100 mV 100 mV to 1 V 1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	0.17 mV/V 0.10 mV/V 0.096 mV /V 0.10 mV/V 0.11 mV/V 23 mV/V

Schedule

Issue date: 06 April 2025
Valid Until: -



NO: SAMM 759

Page: 10 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	1 mV to 100 mV 100 mV to 1 V 1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	0.17 mV/V 0.10 mV/V 0.096 mV /V 0.10 mV/V 0.11 mV/V 23 mV/V
	1 mV to 100 mV 100 mV to 1 V 1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	None
	1 mV to 100 mV 100 mV to 1 V 1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	None
	1 mV to 100 mV 100 mV to 1 V 1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	None
	10 mV to 750 V	See Matrix E
	10 mV to 750 V	None
	750 V to 1000 V	None
Fixed Value	10	97 pa
	1.90	0.18 mQ
	100	0.26 mQ
	190	0.48 mQ
	100 Q	1.3 mQ
	190 O	None
	1kQ	86 mQ
	1.9kQ	23 mQ
	10 kQ	0.120
	19 kQ	0.23 0
	100 kQ	140
	190	2.70
	190	23 0
	190	440
	10 MQ	0.49
	19	0.99 kQ
	100 MQ	17kQ
	10 pF 100 pF	0.01 F/F 0.01 F/F
	10 pF 100 pF	None
Instrument	100 kHz to 26.5 GHz	None
	None	None
	None	None
	None	None
	0.03 to 0.33 mA	None
	1.0 to 33 mV	None
	None	None
	100 yA	None
	None	None
	None	None
	None	None
Measuring Instrument	None	None
	None	None
Resistance	00	40

NO: SAMM 759

Page: 11 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	0 to 100	24 pO/O + 87 pO
	10 QO to 1000	20 + 0.7 ma
	100 © to 1000	12 + 1.9 ma
	1 kQ to 10	12 + 0.2 ma
	10 kQ to 100 kQ	13
	100 kQ to 1000 kQ	22 + 10.0
	1000 kQ to 1 MQ	70 + 192.0
	1 MQ to 10 MO	0.6 + 3.6
	10 to 100	rdg+0.00350
	0.001 Q to 50 O	None
	1 kQ to 10 kQ 10 kQ to 100 kQ	0.13 mQ/Q 0.13 mQ/Q
	100 kQ to 1 MQ	0.13 mQ/Q
	1 MQ to 10 MQ	0.51 mQ/Q
	10 MQ to 100 MQ	9.4mQ/Q
	10 kQ to 100 kQ 100 kQ to 1 MQ	0.13mQ/Q
	1 MQ to 10 MQ	0.47 mQ/Q
	10 MQ to 100 MQ	9.4mQ/Q
	3 Hz to 5 Hz	1.2 mHz/ Hz
Rf Frequency	100 kHz to 26.5 GHz	9.1 pHz/Hz + 70 mHz
	100 kHz to 26.5 GHz	None
	100 kHz to 26.5 GHz	None
	100 kHz to 26.5 GHz	None
	100 kHz to 26.5 GHz	None

SCOPE OF TESTING : ELECTRICAL (RADIO FREQUENCY)

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Ac Current	220 to 2.2A (see Matrix B)	(See Matrix B)
	150 A to 1000 A	None
	1000 Hz:	None
	00to10A 10Ato 100A	1.3 mA/A
	5000 Hz:	None
	00to10A	1.3 mA/A
	10Ato 100A	None
	10000 Hz:	None
	00to10A	1.3 mA/A
	10Ato 100A	None
	See Matrix F	None
	0 to 33 mA 0 to 330 mA 0to22A 0to11A	None
	0.03 to 0.33 mA	None
	10 Hz to 20 Hz	0.3 pA
	20 Hz to 45 Hz	0.3 pA

Schedule

Issue date: 06 April 2025

Valid Until: -



NO: SAMM 759

Page: 12 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	45 Hz to 1 kHz	0.3 pA
	1 kHz to 5 kHz 5 kHz to 10 kHz	0.3 pA 0.3 pA
	0.33 to 3.3 mA	None
	10 Hz to 45 Hz	3 yA
	45 Hz to 1 kHz	3 yA
	1 kHz to 5 kHz	3 yA
	5 kHz to 10 kHz	3 yA
	3.3 to 33 mA	None
	10 Hz to 45 Hz	28
	45 Hz to 1 kHz	28
	1 kHz to 5 kHz	28
	5 kHz to 10 kHz	28
	33 to 330 mA	None
	10 Hz to 45 Hz	0.4mA
	45 Hz to 1 kHz	0.4mA
	1 kHz to 5 kHz	0.4mA
	5 kHz to 10 kHz	0.4mA
	0.33 to 2.2A	None
	10 Hz to 45 Hz	4mA
	45 Hz to 1 kHz	4mA
	1 kHz to 5 kHz	7mA
	2.2to11A	None
	45 Hz to 65 Hz	None
	65 Hz to 500 Hz	13 mA
	500 Hz to 1 KHz	14mA
	10 Hz to 20 Hz 20 Hz to 45 Hz	None
	45 Hz to 1 kHz	None
	1 kHz to 5 kHz	None
	5 kHz to 10 kHz	5 yA
	0.33 to 3.3 mA	None
	10 Hz to 45 Hz	5 yA
	45 Hz to 1 kHz	5 yA
	1 kHz to 5 kHz	None
	5 kHz to 10 kHz	23
	3.3 to 33 mA	None
	10 Hz to 45 Hz	49 pA
	45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	46 pA 83 0.2mA
	33 to 330 mA	None
	10 Hz to 45 Hz	0.6 mA
	45 Hz to 1 kHz	0.5mA
	1 kHz to 5 kHz	0.9mA
	5 kHz to 10 kHz	2.3mA
	0.33 to 2.2A	None
	10 Hz to 45 Hz	6mA
	45 Hz to 1 kHz	5mA
	1 kHz to 5 kHz	20 mA
	2.2to11A	None

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3004176> for the current scope of accreditation

NO: SAMM 759

Page: 13 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	45 Hz to 65 Hz	None
	65 Hz to 500 Hz	20 mA
	500 Hz to 1	45 mA
	33 UA to 10A	See Matrix B
	290 to See Matrix B	See Matrix B
	0.1 20A See Matrix D	See Matrix D
	0.1 UA to 20A See Matrix H	See Matrix H
	See Matrix B	See Matrix B
	See Matrix D	See Matrix D
	100 yA	None
	10 Hz to 1 kHz 1 kHz to 5 kHz	0.0097 pA 0.016 pA
	2.2Ato11A 45 Hz to 65 Hz	1.00mA/A
	65 Hz to 500 Hz	None
	500 Hz to 1 kHz	None
	11 A to 550 A (current coil) 45 Hz to 65 Hz 65 Hz to 440 Hz	3.4mA/A
	100A	None
	1mAto3A	See Matrix D
	1mAto3A	None
	3Ato 100A	None
	45 Hz to 65 Hz	2.4mA/A
	45 Hz to 65 Hz	None
	50 to3A	None
	3Ato 100A	None
	45 Hz to 65 Hz	2.5mA/A
	45 Hz to 65 Hz	None
	45 Hz to 65 Hz	None
Ac Voltage	2.2 mV to 220 V (See Matrix A)	(See Matrix A)
	220 V to 1100 V	None
	15 Hz to 50 Hz	0.30 mV/V + 16 mV
	220 V to 1100 V	None
	50 Hz to 1 kHz	77 + 3.2 mV
	See Matrix E	None
	0.5 kV to 10 kV	12 mV/V + 13 mV
	1.0 to 33 mV	None
	10 Hz to 45 Hz	3.6
	45 Hz to 10 kHz	43
	10 kHz to 20	3.6 UV
	20 kHz to 50 kHz	5.7
	50 kHz to 100 kHz	11
	100 kHz to 500 kHz	26
	33 to 330 mV	None
	10 Hz to 45 Hz	13
	45 Hz to 10 kHz	77 uv
	10 kHz to 20	13
	20 kHz to 50 kHz	18
	50 kHz to 100 kHz 100 kHz to 500 kHz	45 110

NO: SAMM 759

Page: 14 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	0.33 to 3.3 V	None
	10 Hz to 45 Hz	0.1 mV
	45 Hz to 10 kHz	0.1 mV
	10 kHz to 20 kHz	0.1 mV
	20 kHz to 50 kHz	0.2 mV
	50 kHz to 100 kHz	0.2 mV
	100 kHz to 500 kHz	1.1 mV
	3.3 to 33. V	None
	10 Hz to 45 Hz	1.2mV
	45 Hz to 10 kHz	None
	10 kHz to 20 kHz	1.2mV
	20 kHz to 50 kHz	2.1mV
	50 kHz to 100 kHz	2.8 mV
	33 to 330 V	None
	10 Hz to 45 Hz	None
	45 Hz to 10 kHz	14 mV
	10 kHz to 20 kHz	14 mV
	330 to 1020 V	None
	45 Hz to 1 kHz	39 mV
	1 kHz to 5 kHz	52 mV
	-1 A to -100 mA	None
	10 mV rang	None
	0 mV to 10 mV	None
	100 mV rang	None
	10 mV to 100 mV	None
	1V rang	None
	100 mV to 1 V	None
	10 V range	(See Matrix A)
	1Vto10V	None
	100 V rang	None
	10 V to 100 V	None
	1000 V rang	None
	100 V to 700 V	None
	(See Matrix A)	None
	10 Hz to 45 Hz	0.14 mV
	45 Hz to 10 kHz	85
	10 kHz to 20 kHz 20 kHz to 50 kHz	91 0.11 mV
	50 kHz to 100 kHz	0.16 mV
	100 kHz to 500 kHz	0.41 mV
	33 to 330 mV	None
	10 Hz to 45 Hz	0.9 mV
	45 Hz to 10 kHz	0.2 mV
	10 kHz to 20 kHz	None
	20 kHz to 50 kHz	0.6 mV
	50 kHz to 100 kHz	1.0 mV
	100 kHz to 500 kHz	2.8 mV
	0.33 to 3.3 V	None

Schedule

Issue date: 06 April 2025
Valid Until: -



NO: SAMM 759

Page: 15 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	10 Hz to 45 Hz	5.9 mV
	45 Hz to 10 kHz	1.2mV
	10 kHz to 20 kHz	3.1 mV
	20 kHz to 50 kHz	5.6 mV
	50 kHz to 100 kHz	11 mV
	100 kHz to 500 kHz	22 mV
	3.3 to 33 V	None
	10 Hz to 45 Hz	59 mV
	45 Hz to 10 kHz	16 mV
	10 kHz to 20 kHz	33 mV
	20 kHz to 50 kHz	76 mV
	50 kHz to 100 kHz	110 mV
	33 to 330 V	None
	10 Hz to 45 Hz	0.2
	45 Hz to 10 kHz	0.3
	10 kHz to 20 kHz	0.3
	330 to 1020 V	None
	45 Hz to 1 kHz	0.7
	1 kHz to 5 kHz	2.3
	- 1000 V to -100 V	None
	10 mV rang	None
	0 mV to 10 mV	None
	100 mV range	None
	10 mV to 100 mV	None
	1V rang	None
	100 mV to1V	None
	10 V rang	(See Matrix B)
	1Vto10V	None
	100 V rang	None
	10 V to 100 V	None
	1000 V rang	None
	100 V to 700 V	None
	(See Matrix B)	None
	33 mV to 750 V	See Matrix A
	33 mV to 750 V	None
	33 mV to 1020 V See Matrix A	See Matrix A
	100 to 1050 V See Matrix C	See Matrix C
	33 mV to 1020 V See Matrix E	See Matrix E
	100 to 1050 V See Matrix G	See Matrix G
	See Matrix A	See Matrix A
	See Matrix C	See Matrix C
	10 kHz to 20 kHz	1.0mV/V
	330 V to 1020 V	None
	45 Hz to 1 kHz	0.67
	1 kHz to 5 kHz	2.3mV/V
	5 kHz to 10 kHz	2.3mV/V
	1 mV to 750 V	See Matrix C

NO: SAMM 759

Page: 16 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	750 V to 1000 V 45 Hz to 1 kHz 1 kHz to 10 kHz	9.2mV/V 9.2mV/V
	45 Hz to 65 Hz	35
	45 Hz to 65 Hz	None
	45 Hz to 65 Hz	None
	1 mA to 10 mA	0.90 mA/A
	10 mA to 100 mA	0.77
	0.045 kHz to 1 kHz	9.2mV/V
	1 kHz to 10 kHz	9.2mV/V
	to 40 kV 45 Hz to 65 Hz	62 mV/V
	to 40 kV 45 Hz to 65 Hz	None
Dc Current	0 UA to 220 pA	0.94 nA/A + 7.8 nA
	220 UA to 2.2 mA	8.6 nA/A + 24 nA
	2.2 mA to 22 mA	87 + 0.21
	22 mA to 220 mA	1.5 + 3.4
	220 mA to 2.2A	18 + 48 LA
	10 pA	0.2
	10 to 100 pA	37 +
	100 pA to 1mA	30 + 6 nA
	1 mA to 10 mA	30 + 52 nA
	10 mA to 100 mA	47 + 0.6
	100 mA to 1A	0.1 + 5 pA
	0 to 3.3 mA	0.1 pA
	0 to 33 mA 0 to 330 mA 0 to 22A 0 to 11A	2.1 0.1mA 1.2mA 7.8mA
	100 pA	None
	+10 to +100 pA	16 nA
	-100 pA 10 pA	16 nA
	+100 to +1 mA	69 nA
	-1 mA to -100 pA	69 nA
	10 mA	None
	+1 +10 mA	0.7
	-10 mA to -1 mA	0.7 UA
	100 mA	None
	+10 mA to +100 mA	None
	-100 mA to -10 mA	None
	1A	None
	+100 mA to +1A	0.1 mA
	-1 A to -100 mA	0.1 mA
	110 to 330 MQ	(of reading)
	0 to 3.3 mA mA 0 to 330 mA 0 to 22A	0.5 pA 0.1mA 1.4mA
	0 to 3.3 mA mA 0 to 330 mA 0 to 22A	11mA
	+10 to +100 pA -100 pA 10 pA	16 nA 16 nA
	1mA	None
	+100 A to +1 mA	74nA
	mA to -100 pA	74nA

NO: SAMM 759

Page: 17 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	10 mA	None
	+1 mA to +10 mA	0.7 pA
	-10 mA to -1 mA	0.7 pA
	100 mA	None
	+10mA to +100 mA	8 yA
	-100 mA to -10 mA	8 yA
	1A	None
	+100 mMAto+1A	0.2mA
	-1 Ato -100 mA	0.2mA
	0 to 330 pA	170 + 23 nA
	330 UA to 3.3 mA	113 + 57 pA
	3.3 mA to 33 mA	113 + 0.29 PA
	33 mA to 330 mA	113 PA/A + 2.8 LA
	330 mA	227 + 45 pA
	1.1Ato3A	429 + 48 LA
	3Ato10A	567 + 562 PA
	3Ato10A	None
	330 A + (0 pA to 330 A)	+ 0.02
	3.3 mA mA to 3.3 mA)	0.12 + 0.05
	33 mA + (0 mA to 33 mA)	0.12 mA/A PA
	330 mA + (0 mA to 330 mA)	0.12
	1.1A +(0Ato	0.23 mA/A + 47 PA
	3A +(1.1Ato3A)	0.46 mAVA + 0.11 mA
	11A	0.58 mA/A + 0.58 mA
	20.5A +(11A to 20.5 A)	1.2 mA/A + 0.87 mA
	200 (0.1 A to 199.99	13 + 0.6 nA
	2mA (0.2 mA to 199.99 mA)	13 + 0.006 pA
	20 mA (2 mA to 19.999 mA)	16 + 0.06
	200 mA (20 mA to 199.99 mA)	55 + 0.98
	2A (0.2 A to 1.9999 A)	0.21 mA/A + 18
	20A (2 A to 19.999 pA)	0.46 mAVA + 0.46 mA
	330 pA + (0 pA to 330 pA)	0.17 mMA/A + 0.02 pA
	3.3 mA mA to 3.3 mA)	0.12 + 0.05 pA
	33 mA + (0 mA to 33 mA)	0.12 0.3 pA
	330 mA + (0 mA to 330 mA)	0.12
	1.1A +(0Ato	0.23 mA/A + 47 pA
	3A	0.46 + 0.11 mA
	11A	0.58 mA/A + 0.58 mA
	20.5A +(11A to 20.5A)	+ 0.87 mA
	200 pA (0.1 pA to 199.99 yA)	13 nA
	2mA (0.2 mA to 199.99 mA)	13 + 0.006
	(2 mA to 19.999 mA)	16 + 0.06 pA
	200 mA (20 mA to 199.99 mA)	55 + 0.98
	2A (0.2 A to 1.9999 A)	0.21 18 pA
	20A (2 A to 19.999 pA)	0.46 mA/A + 0.46 mA
	0 to 100 pA 0.1 mAto1mA	0.058%of 0.035%of rdg+0.0004mA
	50 to 500 pA	0.06%of rdg+0.21

Schedule

Issue date: 06 April 2025
Valid Until: -



NO: SAMM 759

Page: 18 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	0 to 100 pV 0 to 0 to 10 mV 0 to 100 mV 0 to 1V 0 to 10 V 0 to 100 V 0 to 1000 V	None
	0 to 100 pA 0 to 1mA 0 to 10 mA 0 to 100 mA 0 to 1A	0.013 pA 0.00013 mA 0.0016 mA 0.028 mA 0.000028 A
	0 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 0.330 A 0.330 A 2.2A to 11A	0.17 0.13 0.15 mA/A 0.37 mA/A 0.71 mA/A
	0 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 0.330 A 0.330 A 2.2A to 11A	None
	11 A to 550 A (current coil)	None
	11 A to 550 A (current coil)	None
	0.029 mA to 330 mA	See Matrix B
	0.33 A to 2.2A	None
	10 Hz to 45 Hz	2.5mA/A
	45 Hz to 1 kHz 1 kHz to 5 kHz	None
	100 mA to 1A 1A to 3A	1.3mA/A
	100 mA to 1A 1A to 3A	None
	100A	None
	100A	None
	0 mA to 24 mA	0.95 mA/A
	0 mA to 24 mA	None
	0 mA to 24 mA	12mQ/Q
	10 Q to 100 Q	8.8mQ/Q
	100 mA to 1A 1A to 3A	1.3mA/A 1.6mA/A
	100 mA to 1A 1A to 3A	None
	100A	2.5mA/A
	100A	None
	100A	None
	100A	None
	50 to 3A	See Matrix F
Dc Voltage	0 V to 220 mV	4.2
	220 mV to 2.2 V	6.1 + 1.1
	2.2V to 11V	5.1 + 2.0
	11V 22 V to 220 V 220 V to 1100 V	5.0 uV/V + 3.2 7.4 + 28 8.9 + 0.31 mV
	100 mV	None
	+ 100 to + 100 mV	1.8
	-100 mV to - 100	1.8
	iV	None
	+100 mV to 1V	None
	-1 V to -100 mV	None
	10 V	None
	+1V to +10V	0.1 mV
	-10V to -1V	0.1 mV
	100 V	None
	+10V to +100 V	None

NO: SAMM 759

Page: 19 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	-100 V to -10 V	1mV
	1000 V	None
	+ 100 V to + 1000 V	12 mV
	- 1000 V to -100 V	12 mV
	0 to 330 mV 0to3.3V	30 0.2 mV
	0 to	2.2mV
	33 to 330 V	24 mV
	330 to 1000 V	110 mV
	100 mV	None
	+ 100 to + 100 mV	None
	-100	None
	1V	None
	+100 mV to1V	0.12 mV
	-1 V to -100 mV	0.12 mV
	10V	None
	+1Vto+10V	None
	-10V to-1V	None
	100 V	None
	+10V to +100 V	None
	-100 V to -10 V	None
	1000 V	None
	+ 100 V to + 1000 V	17 mV
	- 1000 V to -100 V	17 mV
	0 to 100 mV	0.004%of rdg+0.004mV
	0 to 50 mV	0.05%of rdg+0.005mV
	30 V to 329 V	0.066 mv /V
	329 V to 1020 V	0.066
	1 mV to 33 V	See Matrix A
	1 mV to 33 V	None
	33 V to 330 V	None
	45 Hz to 1 kHz	0.6 mV/V
	1 kHz to 10 kHz	0.92 mV/V
	1 mV to 100 mV 100 mV to 1V	0.098 mV /V 0.054 0.046 0.059
	1Vto10V 10 V to 100 V 100 V to 1000 V	0.064 mV /V
	1 kV to 30 kV	23 mV/V
	1 kV to 30 kV	None
	1 kV to 30 kV	None
	1 kV to 30 kV	None
	0V to	0.58
	1 mV to 100 mV 100 mV to 1 V	0.17 mV/V 0.10 mV/V 0.096 mV /V
	1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	0.10 mV/V 0.11 mV/V 23 mV/V
	1 mV to 100 mV 100 mV to 1 V	0.17 mV/V 0.10 mV/V 0.096 mV /V
	1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	0.10 mV/V 0.11 mV/V 23 mV/V

NO: SAMM 759

Page: 20 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	1 mV to 100 mV 100 mV to 1 V 1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	None
	1 mV to 100 mV 100 mV to 1 V 1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	None
	1 mV to 100 mV 100 mV to 1 V 1Vto10V 10 V to 100 V 100 V to 1000 V 1 kV to 40 kV	None
	10 mV to 750 V	See Matrix E
	10 mV to 750 V	None
	750 V to 1000 V	None
Fixed Value	10	97 pa
	1.90	0.18 mQ
	100	0.26 mQ
	190	0.48 mQ
	100 Q	1.3 mQ
	190 O	None
	1kQ	86 mQ
	1.9kQ	23 mQ
	10 kQ	0.120
	19 kQ	0.23 0
	100 kQ	140
	190	2.70
	190	23 0
	190	440
	10 MQ	0.49
	19	0.99 kQ
	100 MQ	17kQ
	10 pF 100 pF	0.01 F/F 0.01 F/F
	10 pF 100 pF	None
Instrument	100 kHz to 26.5 GHz	None
	None	None
	None	None
	None	None
	0.03 to 0.33 mA	None
	1.0 to 33 mV	None
	None	None
	100 yA	None
	None	None
	None	None
	None	None
	None	None
Measuring Instrument	None	None
	None	None
Resistance	00	40
	0 to 100	24 pO/O + 87 pO
	10 QO to 1000	20 + 0.7 ma
	100 © to 1000	12 + 1.9 ma

NO: SAMM 759

Page: 21 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	1 kQ to 10	12 + 0.2 ma
	10 kQ to 100 kQ	13
	100 kQ to 1000 kQ	22 + 10.0
	1000 kQ to 1 MQ	70 + 192.0
	1 MQ to 10 MO	0.6 + 3.6
	10 to 100	rdg+0.00350
	0.001 Q to 50 O	None
	1 kQ to 10 kQ 10 kQ to 100 kQ	0.13 mQ/Q 0.13 mQ/Q
	100 kQ to 1 MQ	0.13 mQ/Q
	1 MQ to 10 MQ	0.51 mQ/Q
	10 MQ to 100 MQ	9.4mQ/Q
	10 kQ to 100 kQ 100 kQ to 1 MQ	0.13mQ/Q
	1 MQ to 10 MQ	0.47 mQ/Q
	10 MQ to 100 MQ	9.4mQ/Q
	3 Hz to 5 Hz	1.2 mHz/ Hz
Rf Frequency	100 kHz to 26.5 GHz	9.1 pHz/Hz + 70 mHz
	100 kHz to 26.5 GHz	None
	100 kHz to 26.5 GHz	None
	100 kHz to 26.5 GHz	None
	100 kHz to 26.5 GHz	None

SCOPE OF TESTING : ELECTRICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Am Depth	(+23 dBm to -40 dBm)	0.015 %/%
	fe : 100 to 10 MHz	None
	fm : 20 Hz to 10 kHz	None
	Depth : 5 % to 99 %	None
	fo 10 MHz to 26.5 GHz	0.01 %/%
	fm : 20 Hz to 50 kHz	None
	Depth : 5 % to 99 %	None
	fe : 10 MHz to 26.5 GHz	0.015 %/%
	fm : 50 Hz to 100 kHz	None
	Depth : 5 % to 9 9%	None
Fm Deviation	(+23 dBm to -40 dBm)	9.8 mHz/Hz + 33 mHz
	fe : 100 to 10 MHz	None
	fm :20 Hz to 50 kHz	None
	Af = 20 Hz to 50 kHz	None
	fe: 10 MHz to 1 GHz	9.8 mHz/Hz + 12 mHz
	fm : 20 Hz to 100 kHz Af = 20 Hz to 5 MHz	None

NO: SAMM 759

Page: 22 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	fe: 1 GHz to 26.5 GHz	9.8 mHz/Hz
	fm : 20 Hz to 100 kHz	None
	Af = 20 Hz to 5 MHz	None
	10 MHz to 1 GHz	29 mHz/Hz + 4.1 mHz
	fm : 100 kHz to 5 MHz	None
	Af = 20 Hz to 5 MHz	None
	c: 1 GHz to 26.5 GHz	29 mHz/Hz
	fm : 100 kHz to 5 MHz	None
	Af = 20 Hz to 5 MHz	None
Generating/sourcing	(+20 dBm to -25 dBm)	None
	100 kHz to 26.5 GHz	None
	None	None
Instrument Rf Level	100 kHz to 100MHz 100 MHz to 2.4 GHz	0.05 dB 0.07 dB
	2.4 GHz to 12.4 GHz	0.10 dB
	12.4 GHz to 18 GHz	0.14 dB
	18 GHz to 26.5 GHz	0.16 dB
	(+20 dBm to -120 dBm)	None
	See Matrix C	See Matrix C
	See Matrix C	None
Phase Deviation	(+23 dBm to -40 dBm)	970 +
	fe : 200 kHz to 26.5 GHz	4.2 mrad
	fm : 20 Hz to 5 MHz	None
	Ag = 0.1 rad to 100 rad	None
	Ag = 0.1 rad to 100 rad	None
	Ag = 0.1 rad to 100 rad	None
	Ag = 0.1 rad to 100 rad	None

SCOPE OF TESTING : ELECTRICAL (RADIO FREQUENCY)

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Am Depth	(+23 dBm to -40 dBm)	0.015 %/%
	fe : 100 to 10 MHz	None
	fm : 20 Hz to 10 kHz	None
	Depth : 5 % to 99 %	None
	fo 10 MHz to 26.5 GHz	0.01 %/%
	fm : 20 Hz to 50 kHz	None
	Depth : 5 % to 99 %	None
	fe : 10 MHz to 26.5 GHz	0.015 %/%
	fm : 50 Hz to 100 kHz	None
	Depth : 5 % to 99 %	None
Fm Deviation	(+23 dBm to -40 dBm)	9.8 mHz/Hz + 33 mHz

Schedule

Issue date: 06 April 2025
Valid Until: -



NO: SAMM 759

Page: 23 of 23

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
	fe : 100 to 10 MHz	None
	fm :20 Hz to 50 kHz	None
	Af = 20 Hz to 50 kHz	None
	fe: 10 MHz to 1 GHz	9.8 mHz/Hz + 12 mHz
	fm : 20 Hz to 100 kHz Af = 20 Hz to 5 MHz	None
	fe: 1 GHz to 26.5 GHz	9.8 mHz/Hz
	fm : 20 Hz to 100 kHz	None
	Af = 20 Hz to 5 MHz	None
	10 MHz to 1 GHz	29 mHz/Hz + 4.1 mHz
	fm : 100 kHz to 5 MHz	None
	Af = 20 Hz to 5 MHz	None
	c: 1 GHz to 26.5 GHz	29 mHz/Hz
	fm : 100 kHz to 5 MHz	None
	Af = 20 Hz to 5 MHz	None
Generating/sourcing	(+20 dBm to -25 dBm)	None
	100 kHz to 26.5 GHz	None
	None	None
Instrument Rf Level	100 kHz to 100MHz 100 MHz to 2.4 GHz	0.05 dB 0.07 dB
	2.4 GHz to 12.4 GHz	0.10 dB
	12.4 GHz to 18 GHz	0.14 dB
	18 GHz to 26.5 GHz	0.16 dB
	(+20 dBm to -120 dBm)	None
	See Matrix C	See Matrix C
	See Matrix C	None
Phase Deviation	(+23 dBm to -40 dBm)	970 +
	fe : 200 kHz to 26.5 GHz	4.2 mrad
	fm : 20 Hz to 5 MHz	None
	Ag = 0.1 rad to 100 rad	None
	Ag = 0.1 rad to 100 rad	None
	Ag = 0.1 rad to 100 rad	None
	Ag = 0.1 rad to 100 rad	None

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3004176> for the current scope of accreditation