

# Schedule


Issue date: 12 August 2025  
Valid Until: -



## NO: SAMM 928

(Issue 2, 12 August 2025 replacement  
of SAMM 928 dated 12 August 2025)

Page: 1 of 1

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	Global Instrument (S) Sdn. Bhd. Lot 9, Ground Floor, Grand Industrial Centre, Phase 2, Jalan Bundusan, 88300 Penampang, Kota Kinabalu, Sabah. , 88300, SABAH MALAYSIA
	
<b>ACCREDITED SINCE :</b>	12 AUGUST 2025
<b>FIELD(S) OF CALIBRATION:</b>	FORCE

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

**\* The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

<b>CENTRAL LOCATION</b>	Global Instrument (S) Sdn. Bhd. Lot 9, Ground Floor, Grand Industrial Centre, Phase 2, Jalan Bundusan, 88300 Penampang, Kota Kinabalu, Sabah. , 88300, Sabah
<b>FIELD(S) OF CALIBRATION :</b>	FORCE,

## SCOPE OF CALIBRATION : FORCE

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
Compression Test Machine	0 kN to 2000 kN	3.5 kN	Load Column / Load Cell with reference to EN ISO 7500-1:2018
	0 kN to 2000 kN	None	

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/30044450> for the current scope of accreditation