

Schedule


Issue date: 30 May 2025
Valid Until: -



NO: SAMM 037

(Issue 1, 30 May 2025 replacement
of SAMM 037 dated 30 May 2025)

Page: 1 of 2

LABORATORY LOCATION/ CENTRAL OFFICE:	Hengyuan Refining Company Berhad (Main Laboratory) Batu 1, Jalan Pantai 71000 Port Dickson, Negeri Sembilan , 71000, NEGERI SEMBILAN MALAYSIA
	
ACCREDITED SINCE :	30 MAY 2025
FIELD(S) OF TESTING:	CHEMICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

CENTRAL LOCATION:	Hengyuan Refining Company Berhad (Main Laboratory) Batu 1, Jalan Pantai 71000 Port Dickson, Negeri Sembilan , 71000, Negeri Sembilan
FIELD(S) OF TESTING :	CHEMICAL,

SCOPE OF TESTING : CHEMICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Petroleum And Petroleum Products Ugo/igo Gasoil Jet A-1 Regular Kerosene	Detection of Copper Corrosion from Petroleum Products by Copper Strip Tarnish Test	ASTM D 130
Jet A-1 Tops Naphtha Regular Kerosene	Saybolt Colour of Petroleum Products	ASTM D 156
Ago Gasoil Ugo Gasoil Igo Gasoil	Kinematic Viscosity @ 40 ? C	ASTM D 445
Jet A-1 Regular Kerosene	Smoke Point of Aviation Fuels	ASTM D 1322
Ago Gasoil Igo/ugo Gasoil	ASTM Colour of Petroleum Products	ASTM D 1500
Petroleum And Petroleum Products Liquefied Petroleum Gases (lpg)	Copper Strip Corrosion by LPG Residues in LPG	ASTM D 1838 ASTM D 2158

Schedule

Issue date: 30 May 2025
Valid Until: -



NO: SAMM 037

(Issue 1, 30 May 2025 replacement
of SAMM 037 dated 30 May 2025)

Page: 2 of 2

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Petroleum Products	Density and Relative Density of Liquids by Digital Density Meter Cloud Point of Petroleum Products (optical Detection Stepped Cooling Method) Flash Point by Pensky-Martens Closed Cup Tester	ASTM-D4052 ASTM-D5771 ASTM-D93

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3005081> for the current scope of accreditation