

Schedule


Issue date: 07 July 2025
Valid Until: -



NO: SAMM 1083

(Issue 1, 07 July 2025 replacement
of SAMM 1083 dated 07 July 2025)

Page: 1 of 1

LABORATORY LOCATION/ CENTRAL OFFICE:	ASEAN CT Laboratory, Ong.Com Manufacturing Sdn. Bhd. 34-2, Jalan Tasik Selatan 3, Bandar Tasik Selatan 57000 Kuala Lumpur , 57000, WILAYAH PERSEKUTUAN KUALA LUMPUR MALAYSIA
	
ACCREDITED SINCE :	07 JULY 2025
FIELD(S) OF TESTING:	ELECTRICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

CENTRAL LOCATION:	ASEAN CT Laboratory, Ong.Com Manufacturing Sdn. Bhd. 34-2, Jalan Tasik Selatan 3, Bandar Tasik Selatan 57000 Kuala Lumpur , 57000, Wilayah Persekutuan Kuala Lumpur
FIELD(S) OF TESTING :	ELECTRICAL,

SCOPE OF TESTING : ELECTRICAL

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
Current Transformer (5 To 10000)/5a, (1 To 10000)1a, Burden 1 To 30va, Power Factor 0.8 To 1.0 Accuracy Class 0.1, 0.2, 0.5, 1, 3, 5, 0.2s And 0.5s Class	Accuracy Test 1. Ratio Error 2. Phase displacement error	IEC 61869-2:2012 clause 7.3.5

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3005599> for the current scope of accreditation