


NO: SAMM 704

Page: 1 of 1

| | |
|---|---|
| LABORATORY LOCATION/ CENTRAL OFFICE: | Yanmar Kota Kinabalu R&D Center Sdn. Bhd. Lot 11 & 12, IZ4, Road KKIP 1G, South KKIP Kota Kinabalu Industrial Park, 88450 Kota Kinabalu, Sabah, 88450, SABAH MALAYSIA |
|  | |
| ACCREDITED SINCE : | 31 JULY 2025 |
| FIELD(S) OF TESTING: | CHEMICAL |

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

| | |
|------------------------------|---|
| CENTRAL LOCATION: | Yanmar Kota Kinabalu R&D Center Sdn. Bhd. Lot 11 & 12, IZ4, Road KKIP 1G, South KKIP Kota Kinabalu Industrial Park, 88450 Kota Kinabalu, Sabah, 88450, Sabah |
| FIELD(S) OF TESTING : | CHEMICAL, |

SCOPE OF TESTING : CHEMICAL

| Material / Product Tested | Type Of Test / Properties Measured / Range Of Measurement | Standard Test Methods / Equipment / Techniques |
|---|--|--|
| Petroleum Products -biodiesel -biodiesel Blend -diesel | Acid Number Density Flash Point | In-house method TM-01 Rev. 1 based on ASTM D664-18e2 ASTM D4052-22 ASTM D93-20 (Procedure A & C) |
| Petroleum Products -lubricant | Density Acid Number Flash Point Based Number Kinematic Viscosity @ 40°C to 100°C | ASTM D4052-22 In-house method TM-01 Rev. 1 based on ASTM D664-18e2 ASTM D93-20 (Procedure A) ASTM D4739 ASTM D7042 |