


NO: SAMM 873

Page: 1 of 3

|   |   |
|---|---|
| <b>LABORATORY LOCATION/<br/>CENTRAL OFFICE:</b>                                   | APM Calibration Centre, APM Nuclear Technology Sdn. Bhd.<br>No.6, Suntract Hub Bangi, Off Jln P1a, Section 13, Bandar Baru<br>Bangi, 43000 Kajang, Selangor Darul Ehsan, Malaysia. , 43000,<br>SELANGOR<br>MALAYSIA |
|  |   |
| <b>ACCREDITED SINCE :</b>   | 20 AUGUST 2025  |
| <b>FIELD(S) OF CALIBRATION:</b>   | RADIOACTIVITY   |

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

**\* The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

|                                  |   |
|----------------------------------|---|
| <b>CENTRAL LOCATION</b>          | APM Calibration Centre, APM Nuclear Technology Sdn. Bhd.<br>No.6, Suntract Hub Bangi, Off Jln P1a, Section 13, Bandar Baru<br>Bangi, 43000 Kajang, Selangor Darul Ehsan, Malaysia. , 43000,<br>Selangor |
| <b>FIELD(S) OF CALIBRATION :</b> | RADIOACTIVITY,  |

**SCOPE OF CALIBRATION : RADIOACTIVITY**

| <b>Instrument<br/>Calibrated/Measurement<br/>Parameter</b> | <b>Range</b>  | <b>Calibration and<br/>Measurement<br/>Capability<br/>Expressed as an<br/>Uncertainty (<math>\pm</math>)*</b> | <b>Remarks</b>  |
|--|---|---|---|
| Survey Meter / Air Kerma Rate. Ka                          | X-ray (40 kV to 300<br>kV)<br>8.6 $\mu$ Gy/h to 21.1<br>mGy/h | 1.9 % of reading  | Reference from<br>SAFETY REPORT<br>SERIES No. 16<br>Calibration of radiation<br>protection monitoring<br>instruments, IAEA<br>(2000), International<br>Organization of<br>Standard (ISO) 403 –<br>1 and ISO 4037– 3 |

# Schedule

Issue date: 20 August 2025  
Valid Until: -



NO: SAMM 873

Page: 2 of 3

| Instrument<br>Calibrated/Measurement<br>Parameter                          | Range   | Calibration and<br>Measurement<br>Capability<br>Expressed as an<br>Uncertainty ( $\pm$ )* | Remarks   |
|--|---|---|---|
|  | $\text{Æ}''$ – ray (Cs – 137)<br>3.5 $\mu\text{Gy/h}$ to 19.1<br>$\text{mGy/h}$ | 2.6 % of reading  | Reference from<br>SAFETY REPORT<br>SERIES No. 16<br>Calibration of radiation<br>protection monitoring<br>instruments, IAEA<br>(2000), International<br>Organization of<br>Standard (ISO) 403 –<br>1 and ISO 4037– 3 |
| Survey Meter / Ambient Dose<br>Equivalent Rate, $\text{H}^*$ (10)          | X-ray (40 kV to 300<br>kV)<br>14.7 $\mu\text{Gy/h}$ to 34.1<br>$\text{mGy/h}$   | 4.4 % of reading  | Reference from<br>SAFETY REPORT<br>SERIES No. 16<br>Calibration of radiation<br>protection monitoring<br>instruments, IAEA<br>(2000), International<br>Organization of<br>Standard (ISO) 403 –<br>1 and ISO 4037– 3 |
|  | $\text{Æ}''$ – ray (Cs – 137)<br>4.2 $\mu\text{Gy/h}$ to 22.8<br>$\text{mGy/h}$ | 4.8 % of reading  | Reference from<br>SAFETY REPORT<br>SERIES No. 16<br>Calibration of radiation<br>protection monitoring<br>instruments, IAEA<br>(2000), International<br>Organization of<br>Standard (ISO) 403 –<br>1 and ISO 4037– 3 |
| Personal Dosimeter / Personal<br>Dose Equivalent, $\text{H}_\text{p}$ (10) | X-ray (40 kV to 300<br>kV)<br>14.7 $\mu\text{Gy/h}$ to 34.1<br>$\text{mGy/h}$   | 4.3 % of reading  | Reference from<br>SAFETY REPORT<br>SERIES No. 16<br>Calibration of radiation<br>protection monitoring<br>instruments, IAEA<br>(2000), International<br>Organization of<br>Standard (ISO) 403 –<br>1 and ISO 4037– 3 |

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3006274> for the current scope of accreditation

## Schedule

Issue date: 20 August 2025  
Valid Until: -



NO: SAMM 873

Page: 3 of 3

| Instrument<br>Calibrated/Measurement<br>Parameter | Range                                  | Calibration and<br>Measurement<br>Capability<br>Expressed as an<br>Uncertainty ( $\pm$ )* | Remarks   |
|---|--|---|---|
|   | $\text{Æ}''$ – ray (Cs – 137)<br>1 mSv | 4.3 % of reading  | Reference from<br>SAFETY REPORT<br>SERIES No. 16<br>Calibration of radiation<br>protection monitoring<br>instruments, IAEA<br>(2000), International<br>Organization of<br>Standard (ISO) 403 –<br>1 and ISO 4037– 3 |

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3006274> for the current scope of accreditation