

# Schedule

Issue date: 20 August 2025  
Valid Until: -



NO: SAMM 873

Page: 1 of 3

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	APM Calibration Centre, APM Nuclear Technology Sdn. Bhd. No.6, Suntract Hub Bangi, Off Jln P1a, Section 13, Bandar Baru Bangi, 43000 Kajang, Selangor Darul Ehsan, Malaysia. , 43000, SELANGOR MALAYSIA
	
<b>ACCREDITED SINCE :</b>	20 AUGUST 2025
<b>FIELD(S) OF CALIBRATION:</b>	RADIOACTIVITY

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

**\* The uncertainty covered by the CMC is expressed as the expanded uncertainty corresponding to a coverage probability of approximately 95 % and have a coverage factor of k=2 unless stated otherwise.**

<b>CENTRAL LOCATION</b>	APM Calibration Centre, APM Nuclear Technology Sdn. Bhd. No.6, Suntract Hub Bangi, Off Jln P1a, Section 13, Bandar Baru Bangi, 43000 Kajang, Selangor Darul Ehsan, Malaysia. , 43000, Selangor
<b>FIELD(S) OF CALIBRATION :</b>	RADIOACTIVITY,

## SCOPE OF CALIBRATION : RADIOACTIVITY

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
Survey Meter / Air Kerma Rate. Ka	X-ray (40 kV to 300 kV) 8.6 $\mu$ Gy/h to 21.1 mGy/h	1.9 % of reading	Reference from SAFETY REPORT SERIES No. 16 Calibration of radiation protection monitoring instruments, IAEA (2000), International Organization of Standard (ISO) 403 – 1 and ISO 4037– 3

# Schedule

Issue date: 20 August 2025  
Valid Until: -



NO: SAMM 873

Page: 2 of 3

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
	$\text{Ae}^{\text{m}}$ – ray (Cs – 137) 3.5 $\mu\text{Gy/h}$ to 19.1 $\text{mGy/h}$	2.6 % of reading	Reference from SAFETY REPORT SERIES No. 16 Calibration of radiation protection monitoring instruments, IAEA (2000), International Organization of Standard (ISO) 403 – 1 and ISO 4037– 3
Survey Meter / Ambient Dose Equivalent Rate, H* (10)	X-ray (40 kV to 300 kV) 14.7 $\mu\text{Gy/h}$ to 34.1 $\text{mGy/h}$	4.4 % of reading	Reference from SAFETY REPORT SERIES No. 16 Calibration of radiation protection monitoring instruments, IAEA (2000), International Organization of Standard (ISO) 403 – 1 and ISO 4037– 3
	$\text{Ae}^{\text{m}}$ – ray (Cs – 137) 4.2 $\mu\text{Gy/h}$ to 22.8 $\text{mGy/h}$	4.8 % of reading	Reference from SAFETY REPORT SERIES No. 16 Calibration of radiation protection monitoring instruments, IAEA (2000), International Organization of Standard (ISO) 403 – 1 and ISO 4037– 3
Personal Dosimeter / Personal Dose Equivalent, Hp (10)	X-ray (40 kV to 300 kV) 14.7 $\mu\text{Gy/h}$ to 34.1 $\text{mGy/h}$	4.3 % of reading	Reference from SAFETY REPORT SERIES No. 16 Calibration of radiation protection monitoring instruments, IAEA (2000), International Organization of Standard (ISO) 403 – 1 and ISO 4037– 3

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3006274> for the current scope of accreditation

# Schedule

Issue date: 20 August 2025  
Valid Until: -



NO: SAMM 873

Page: 3 of 3

Instrument Calibrated/Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty ( $\pm$ )*	Remarks
	$\text{Æ}^{\text{r}}$ – ray (Cs – 137) 1 mSv	4.3 % of reading	Reference from SAFETY REPORT SERIES No. 16 Calibration of radiation protection monitoring instruments, IAEA (2000), International Organization of Standard (ISO) 403 – 1 and ISO 4037– 3

Scan this QR Code or visit <https://accreditation.ism.gov.my/public/listing/cab/samm-ct/3006274> for the current scope of accreditation