

NO: SAMM 1188

Page: 1 of 1

<b>LABORATORY LOCATION/ CENTRAL OFFICE:</b>	Transformer Testing & Research Centre Transresin Sdn. Bhd. No. 50, Jalan Industri USJ 1/1, Taman Perindustrian USJ 1, Subang Jaya , 47600, SELANGOR MALAYSIA
	
<b>ACCREDITED SINCE :</b>	21 AUGUST 2025
<b>FIELD(S) OF TESTING:</b>	ELECTRICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

<b>CENTRAL LOCATION:</b>	Transformer Testing & Research Centre Transresin Sdn. Bhd. No. 50, Jalan Industri USJ 1/1, Taman Perindustrian USJ 1, Subang Jaya , 47600, Selangor
<b>FIELD(S) OF TESTING :</b>	ELECTRICAL,

**SCOPE OF TESTING : ELECTRICAL**

Material / Product Tested	Type Of Test / Properties Measured / Range Of Measurement	Standard Test Methods / Equipment / Techniques
<b>current Transformer</b> (50 To 2000)/5 A, burden 1.875, 3.75, 7.5 And 15 Va power Factor 0.8 & 1.0 accuracy Class 0.1, 0.2, 0.5, 1.0, 0.2s And 0.5s	<b>Accuracy Test</b>  1. Ratio Error 2. Phase Displacement	IEC 61869-2 : 2012 Clause 7.3.5